Triple balloon enteroscopy: a combined method

Double-balloon enteroscopy (DBE) has high diagnostic and therapeutic yield (70%–85%) [1], but low completion rates (18%–66%) [2,3]. The NaviAid device (SMART Medical Systems Ltd., Ra'anana, Israel) consists in a through-the-scope (TTS) balloon for deep enteroscopy using a colonoscope [4]. In this case, we combined both types of enteroscopy to achieve deeper insertion.

A 52-year-old man was admitted for melena with negative gastroscopy and colonoscopy. Capsule endoscopy showed angioectasia in the jejunum and ileum (► **Fig. 1**).

Before the enteroscopy began, the TTS balloon (NaviAid AB) was introduced through the 3.2-mm working channel of an EN-580T double-balloon enteroscope (Fujinon Inc., Japan). An oral DBE was performed advancing 3.5 m from the ligament of Treitz. After five ineffective progression attempts, the TTS balloon was pushed forward 20–30 cm out of the enteroscope. The TTS balloon was inflated and anchored in the small bowel. The enteroscope balloon was deflated and the enteroscope was pushed as far as the TTS balloon, before the enteroscope balloon was again inflated. The overtube balloon was then deflated and the overtube was pushed to the distal section of the enteroscope. Once the three inflated balloons were together, they were pulled back under fluoroscopic guidance. The same steps were performed repeatedly, achieving a further advance of 1.5 m (► **Video 1**). During the retrieval, angioectasias were treated with argon plasma coagulation.

Endoscopy_UCTN_Code_TTT_1AP

Competing interests

None

► **Fig. 1** Capsule endoscopy images showing angioectasias in the small bowel.

► **Video 1** Schematic showing the technique of triple-balloon enteroscopy, along with views during the procedure which allowed for improved advancement of the enteroscope through the small bowel.

The authors

Gerardo Blanco-Velasco, Raúl Antonio Zamarripa-Mottu, Omar Michel Solórzano-Pineda, Oscar Víctor Hernández-Mondragón, Juan Manuel Blancas-Valencia
Department of Endoscopy, Hospital de Especialidades, Centro Médico Nacional Siglo XXI, Instituto Mexicano del Seguro Social, Mexico City, Mexico

Corresponding author

Gerardo Blanco-Velasco, MD
Department of Endoscopy, Hospital de Especialidades, Centro Médico Nacional Siglo XXI, Instituto Mexicano del Seguro Social, Av. Cuauhtémoc 330, Mexico City, 06725, Mexico
Fax: +52-55-740535
gerardoblancov@hotmail.com

References

- [1] May A. Double-balloon enteroscopy. *Gastrointest Endoscopy Clin N Am* 2017; 27: 113–122
- [2] Domagk D, Mensink P, Aktas H et al. Single vs. double-balloon enteroscopy in small-bowel diagnostics: a randomized multicenter trial. *Endoscopy* 2011; 43: 472–476
- [3] May A, Farber M, Aschmoneit I et al. Prospective multicenter trial comparing push-and-pull enteroscopy with the single- and double-balloon techniques in patients with small-bowel disorders. *Am J Gastroenterol* 2010; 105: 575–581
- [4] Ali R, Wild D, Shieh F et al. Deep enteroscopy with a conventional colonoscope: initial multicenter study by using a through-the-scope balloon catheter system. *Gastrointest Endosc* 2015; 82: 855–860

Bibliography

DOI <https://doi.org/10.1055/a-0966-8431>
Published online: 24.7.2019
Endoscopy 2019; 51: E390–E391
© Georg Thieme Verlag KG
Stuttgart · New York
ISSN 0013-726X

ENDOSCOPY E-VIDEOS

<https://eref.thieme.de/e-videos>

Endoscopy E-Videos is a free access online section, reporting on interesting cases and new

techniques in gastroenterological endoscopy. All papers include a high quality video and all contributions are freely accessible online.

This section has its own submission website at

<https://mc.manuscriptcentral.com/e-videos>