Percutaneous endoscopic removal of a biliary metal stent retained in the jejunum using a digital cholangioscope

Temporary placement of a biliary metal stent is a useful percutaneous treatment for benign hepaticojunal anastomotic stricture [1–4]. However, a metal stent that has been placed will sometimes migrate spontaneously and can be retained in the jejunum. Endoscopic removal of a migrated metal stent retained in the jejunum is sometimes technically challenging and, in difficult cases, invasive surgical removal is required. Here, we describe the successful non-invasive endoscopic removal of a retained metal stent using a digital cholangioscope that was passed via the percutaneous fistula created by a percutaneous transhepatic biliary drain. A 70-year-old man with a surgical hepaticojunostomy had suffered from recurrent cholangitis due to hepaticojunal anastomotic stricture. As an endoscopic approach to the anastomosis using a balloon enteroscope failed because of severe adhesions, percutaneous transhepatic biliary drainage was performed; however, frequent drainage catheter replacement was required. A fully covered biliary metal stent with a lasso (HANARO-STENT Biliary; Boston Scientific Corp.) was therefore placed at the anastomosis via the percutaneous fistula (► Fig. 1). The metal stent spontaneously migrated several days later and was retained in a sharp bend in the jejunum near to the anastomosis for 4 months, although the anastomotic stricture had improved (► Fig. 2). The patient refused surgical removal of the stent, so endoscopic removal via the percutaneous fistula using a digital cholangioscope was attempted (► Video 1).

First, a 0.025-inch guidewire was placed near the retained metal stent. A digital cholangioscope (SpyGlass DS System;

► **Fig. 1** Radiographic images of the percutaneous temporary metal stent placement showing: **a** the severe hepaticojunal anastomotic stricture (arrow); **b** a fully covered biliary metal stent with a lasso in position at the anastomosis.

Boston Scientific Corp.) was then inserted using guidewire guidance into the jejunum through the hepaticojunal anastomosis via the percutaneous fistula. The lasso of the metal stent was grasped with a biopsy forceps under direct visualization and was retracted into the cholangioscope. The cholangioscope was gradually pulled out through the intrahepatic bile duct and the retained metal stent was successfully removed via the percutaneous fistula. This technique is a useful and non-invasive option for removal of a retained stent.

Endoscopy_UCTN_Code_CPL_1AK_2AD

Competing interests

T. Itoi has received speaker's fees from Boston Scientific Japan. The remaining authors declare that they have no conflict of interest.

► **Fig. 2** Radiographic image showing the metal stent (arrow) that had spontaneously migrated 4 months previously and was retained in the jejunum.

Video 1 Endoscopic removal of a biliary metal stent retained in the jejunum was performed using a digital cholangioscope via the percutaneous fistula.

Bibliography

Endoscopy 2021; 53: E201–E202
DOI 10.1055/a-1234-6062
ISSN 0013-726X
 published online 11.9.2020
 © 2020. Thieme. All rights reserved.
 Georg Thieme Verlag KG, Rüdigerstraße 14,
 70469 Stuttgart, Germany

ENDOSCOPY E-VIDEOS

<https://eref.thieme.de/e-videos>

Endoscopy E-Videos is a free access online section, reporting on interesting cases and new

techniques in gastroenterological endoscopy. All papers include a high quality video and all contributions are freely accessible online.

This section has its own submission website at
<https://mc.manuscriptcentral.com/e-videos>

The authors

Shuntaro Mukai, Takao Itoi, Atsushi Sofuni, Takayoshi Tsuchiya, Kentaro Ishii, Reina Tanaka, Ryosuke Tonoza

Department of Gastroenterology and Hepatology, Tokyo Medical University, Tokyo, Japan

Corresponding author

Takao Itoi, MD, PhD

Department of Gastroenterology and Hepatology, Tokyo Medical University,
 6-7-1 Nishishinjuku, Shinjuku-ku,
 Tokyo 160-0023, Japan
 Fax: +81-3-53816654
 itoi@tokyo-med.ac.jp

References

- [1] Yun G, Yoon CJ, Seong NJ. Percutaneous treatment of benign bilioenteric anastomotic strictures: temporary covered stent placement versus balloon dilatation. *Eur Radiol* 2019; 29: 2690–2697
- [2] Hai R, Kuban J. Percutaneous intervention for bilioenteric anastomotic strictures: current strategies and future directions. *Gastrointest Interv* 2017; 6: 70–77
- [3] Walter D, Laleman W, Jansen JM et al. A fully covered self-expandable metal stent with antimigration features for benign biliary strictures: a prospective, multicenter cohort study. *Gastrointest Endosc* 2015; 81: 1197–1203
- [4] Gwon DI, Ko GY, Ko HK et al. Percutaneous transhepatic treatment using retrievable covered stents in patients with benign biliary strictures: mid-term outcomes in 68 patients. *Dig Dis Sci* 2013; 58: 3270–3279