Endoscopic ultrasound-guided choledochoduodenostomy without fistula dilation using a novel fully covered metallic stent with a 5.9-Fr ultra-thin delivery system

Endoscopic ultrasonography-guided choledochoduodenostomy (EUS-CDS) is a potential procedure for primary drainage in unresectable malignant distal biliary obstruction and can replace endoscopic retrograde cholangiopancreatography. However, adverse events following EUS-CDS are occasionally reported, the most common being bile leak due to fistula dilation [1,2]. The safety of EUS-CDS without fistula dilation using two types of thin, fully covered self-expandable metallic stents (FCSEMSs) has been previously reported [3–5]. Here we report a case in which EUS-CDS without fistula dilation was successfully performed using a novel FCSEMS with a 5.9-Fr delivery system (Hanarostent Benefit; M.I.Tech, Seoul, Korea) (► **Fig. 1**).

A 62-year-old man who had received chemotherapy for unresectable pancreatic body cancer developed a distal biliary obstruction. Computed tomography revealed a 25-mm mass on the pancreatic body and a dilated common hepatic duct (CHD) (► **Fig. 2a**). We performed EUS-CDS for primary drainage.

The dilated CHD was localized using a forward-viewing echoendoscope (TGF-UC260J; Olympus Medical Systems, Tokyo, Japan) from the duodenal bulb (► **Fig. 2b**). First, the CHD was punctured with a 19-gauge needle (EZ Shot 3 Plus; Olympus Medical Systems). Cholangiography revealed a dilated CHD with distal obstruction. Second, a 0.025-inch guidewire (M-Through; ASAHI INTECC Corp., Tokyo, Japan) was inserted into the B4 branch. Fistula dilation was avoided, and the novel FCSEMS (8 mm × 6 cm) was passed through the duodenum and CHD wall smoothly. Finally, the stent was placed in the CHD from the duodenal bulb (► **Fig. 3**, ► **Fig. 4**; ► **Video 1**).

No adverse events occurred during or after the procedure.

► **Fig. 1** Novel fully covered self-expandable metal stent with an ultra-thin delivery system. **a** The outer sheath of the delivery catheter is size 5.9Fr. **b** The expanded stent with a braiding design.

► **Fig. 2** Image findings before the procedure. **a** Coronal computed tomographic image revealing a hypovascular mass of the pancreatic body (arrows) and dilated common hepatic duct (CHD) (arrowhead). **b** Endoscopic ultrasonography image indicating a hypoechoic mass of the pancreatic body (arrows) and dilated CHD (arrowhead).

In previous reports, EUS-CDS without fistula dilation was performed using the FCSEMSs with a 7-Fr or 7.5-Fr delivery system, with 31.6%–100% technical success rate [3–5]. This is the first report discussing EUS-CDS without fistula dilation using a FCSEMS with a 5.9-Fr delivery system, which is the thinnest ever commercially available delivery system and is capable of simpler and safer EUS-CDS procedures.

Endoscopy_UCTN_Code_TTT_1AS_2AD

Acknowledgements

This work was supported in part by The National Cancer Center Research and Development Fund (31-A-13).

Competing interests

The authors declare that they have no conflict of interest.

► **Fig. 3** Endoscopic ultrasonography-guided choledochoduodenostomy without fistula dilation. **a** The common hepatic duct (CHD) was punctured with a 19-gauge needle. **b** Cholangiography revealed a dilated CHD with a distal obstruction. **c** A 0.025-inch guidewire was inserted into the B4 branch, and the fully covered self-expandable metal stent (FCSEMS) was inserted into the CHD without fistula dilation. **d** The FCSEMS was deployed into the CHD.

► **Fig. 4** Image findings after the procedure. **a** Fluoroscopic image. **b** Endoscopic image.

The authors

Takehiko Koga¹, Susumu Hijioka¹, Yuya Hisada¹, Yuta Maruki¹, Yoshikuni Nagashio¹, Takuji Okusaka¹, Yutaka Saito²

1 Department of Hepatobiliary and Pancreatic Oncology, National Cancer Center Hospital, Tokyo, Japan

2 Department of Endoscopy, Gastrointestinal Endoscopy Division, National Cancer Center Hospital, Tokyo, Japan

Corresponding author

Susumu Hijioka

Department of Hepatobiliary and Pancreatic Oncology, National Cancer Center Hospital, 5-1-1 Tsukiji, Chuo-ku, Tokyo 104-0045, Japan

Fax: +81-3-3542-3815

shijioka@ncc.go.jp

References

- [1] Kawakubo K, Isayama H, Kato H et al. Multi-center retrospective study of endoscopic ultrasound-guided biliary drainage for malignant biliary obstruction in Japan. *J Hepatobiliary Pancreat Sci* 2014; 21: 328–334
- [2] Wang K, Zhu J, Xing L et al. Assessment of efficacy and safety of EUS-guided biliary drainage: a systematic review. *Gastrointest Endosc* 2016; 83: 1218–1227
- [3] Park DH, Lee TH, Paik WH et al. Feasibility and safety of a novel dedicated device for one-step EUS-guided biliary drainage: a randomized trial. *J Gastroenterol Hepatol* 2015; 30: 1461–1466
- [4] Itonaga M, Kitano M, Hatamaru K et al. Endoscopic ultrasound-guided choledochoduodenostomy using a thin stent delivery system in patients with unresectable malignant distal biliary obstruction: a prospective multicenter study. *Dig Endosc* 2019; 31: 291–298
- [5] Paik WH, Lee TH, Park DH et al. EUS-guided biliary drainage versus ERCP for the primary palliation of malignant biliary obstruction: a multicenter randomized clinical trial. *Am J Gastroenterol* 2018; 113: 987–997

Video 1 The video shows an endoscopic ultrasonography-guided choledochoduodenostomy without fistula dilation performed using a novel fully covered self-expandable metal stent with a 5.9-Fr ultra-thin delivery system.

Bibliography

Endoscopy 2021; 53: E223–E225

DOI 10.1055/a-1244-9651

ISSN 0013-726X

published online 11.9.2020

© 2020. Thieme. All rights reserved.

Georg Thieme Verlag KG, Rüdigerstraße 14,
70469 Stuttgart, Germany

ENDOSCOPY E-VIDEOS

<https://eref.thieme.de/e-videos>

Endoscopy E-Videos is a free access online section, reporting on interesting cases and new techniques in gastroenterological endoscopy. All papers include a high quality video and all contributions are freely accessible online.

This section has its own submission website at

<https://mc.manuscriptcentral.com/e-videos>