Small-caliber plastic stent for endoscopic ultrasound-guided drainage of a non-dilated pancreatic duct

Endoscopic ultrasound (EUS)-guided pancreatic drainage is increasingly utilized in cases where endoscopic retrograde cholangiopancreatography (ERCP) is not possible [1–3]. EUS-guided pancreatic drainage for a non-dilated pancreatic duct (PD) not only poses technical challenges but also increases the risk of procedure-related pancreatitis as well as stent-induced ductal injury. A small-caliber stent might be suitable for a non-dilated PD to prevent ductal injury and upstream main duct obstruction, but there has been no report on a dedicated 5-Fr EUS-guided pancreatic drainage stent. Herein we present a newly designed 5-Fr plastic stent for EUS-guided pancreatic drainage (Through & Pass Type IT; Gadelius Medical, Tokyo, Japan) (►Fig. 1), which is a thinner type of a previously reported stent [4].

A 75-year-old woman with a history of pancreatoduodenectomy for intraductal papillary mucinous neoplasm was hospitalized with a pancreatic fluid collection (►Fig. 2). The fluid collection did not subside after percutaneous drainage, and therefore we decided to perform EUS-guided pancreatic drainage (►Video 1). Under EUS guidance, we punctured the non-dilated main PD with a 19-gauge needle and inserted a 0.025-inch guidewire (►Fig. 3a). The needle tract was dilated using an ultra-tapered mechanical dilator (ES dilator; Zeon Medical, Tokyo, Japan) and a 4-mm-wide balloon dilator (REN; Kaneka, Osaka, Japan). Using a double-lumen catheter (Uneven Double Lumen Cannula; Piolax Medical Devices, Kanagawa, Japan), we successfully passed a 0.025-inch hydrophilic guidewire into the jejunum and additionally inserted a 0.035-inch guidewire to stabilize the scope position (“double guidewire technique”) (►Fig. 3b) [5]. After dilation of the pancreaticojejunostomy anastomosis using the balloon dilator, a 5-Fr stent was readily positioned from the jejunum to the stomach

►Fig. 1 Newly designed 5-Fr pigtail plastic stent for endoscopic ultrasound-guided pancreatic drainage. The stent has a pigtail structure at one end and two flanges at each end to prevent migration. This stent was developed as a modification of the 7-Fr plastic stent designed for endoscopic ultrasound-guided drainage.

►Fig. 2 Contrast-enhanced computed tomography showing a fluid collection around the pancreaticojejunostomy anastomosis (arrow). An endovascular stent for celiac artery stenosis (arrowhead) is also shown.

(►Fig. 3c). The postprocedural course was uneventful and the percutaneous drain was successfully removed. A follow-up CT scan 4 months later revealed the disappearance of the fluid collection without upstream ductal dilation. Our new 5-Fr plastic stent was feasible in the EUS-guided drainage of a non-dilated PD.

Endoscopy_UCTN_Code_TTT_1AS_2AD

Competing interests

Dr. Itoi serves as a consultant of Gadelius Medical.

►Fig. 3 Endoscopic ultrasound-guided transmurals placement of a 5-Fr plastic stent for the non-dilated pancreatic duct. **a** Endoscopic ultrasound-guided access to the main pancreatic duct using a 19-gauge needle. Pancreatography delineated the non-dilated pancreatic duct with a diameter of 1.6 mm. A percutaneous catheter is also shown. **b** We passed a 0.025-inch guidewire into the jejunum. Using a double-lumen catheter, we additionally inserted a 0.035-inch guidewire, which was utilized to stabilize the scope position (“double guidewire technique”). **c** Placement of a 5-Fr stent across the pancreaticojejunostomy anastomotic stricture.

Video 1 Endoscopic ultrasound-guided transmural placement of a 5-Fr plastic stent for a non-dilated pancreatic duct.

Bibliography

Endoscopy 2021; 53: E407–E408
DOI 10.1055/a-1333-0538
ISSN 0013-726X
published online 14.1.2021
 © 2021. Thieme. All rights reserved.
 Georg Thieme Verlag KG, Rüdigerstraße 14,
 70469 Stuttgart, Germany

ENDOSCOPY E-VIDEOS

<https://eref.thieme.de/e-videos>

Endoscopy E-Videos is a free access online section, reporting on interesting cases and new

techniques in gastroenterological endoscopy. All papers include a high quality video and all contributions are freely accessible online.

This section has its own submission website at
<https://mc.manuscriptcentral.com/e-videos>

The authors

**Dosuke Iwadate¹, Yousuke Nakai^{1,2},
 Tsuyoshi Hamada¹, Takao Itoi³, Kazuhiko Koike¹**

- 1 Department of Gastroenterology, Graduate School of Medicine, The University of Tokyo, Tokyo, Japan
- 2 Department of Endoscopy and Endoscopic Surgery, Graduate School of Medicine, The University of Tokyo, Tokyo, Japan
- 3 Department of Gastroenterology and Hepatology, Tokyo Medical University, Tokyo, Japan

Corresponding author

Yousuke Nakai, MD

Department of Endoscopy and Endoscopic Surgery, Graduate School of Medicine, The University of Tokyo, 7-3-1 Hongo, Bunkyo-ku, Tokyo 113-8655, Japan
 ynakai-tky@umin.ac.jp

References

- [1] Kahaleh M, Artifon ELA, Perez-Miranda M et al. EUS-guided drainage: summary of therapeutic EUS consortium meeting. *Endosc Ultrasound* 2019; 8: 151–160
- [2] Teoh AYB, Dhir V, Kida M et al. Consensus guidelines on the optimal management in interventional EUS procedures: results from the Asian EUS group RAND/UCLA expert panel. *Gut* 2018; 67: 1209–1228
- [3] Itoi T, Kasuya K, Sofuni A et al. Endoscopic ultrasonography-guided pancreatic duct access: techniques and literature review of pancreatography, transmural drainage and rendezvous techniques. *Dig Endosc* 2013; 25: 241–252
- [4] Matsunami Y, Itoi T, Sofuni A et al. Evaluation of a new stent for EUS-guided pancreatic duct drainage: long-term follow-up outcome. *Endosc Int Open* 2018; 6: E505–E512
- [5] Nakai Y, Kogure H, Koike K. Double-guide-wire technique for endoscopic ultrasound-guided pancreatic duct drainage. *Dig Endosc* 2019; 31: 65–66