Hemorrhage after laser-cut covered self-expandable metal stent removal

Endoscopic covered self-expandable metal stent (CSEMS) placement is widely used for distal malignant biliary obstruction (MBO) as the stent prevents ingrowth and is easily removed [1,2]. The usefulness of the laser-cut CSEMS (LC-CSEMS) and its endoscopic removal have been reported [3–5]; there have been no previous reports of complications related to LC-CSEMS removal. We here report a case of hemorrhage following LC-CSEMS removal.

An 83-year-old woman was referred for cholangitis. She had undergone endoscopic LC-CSEMS (X-Suit NIR covered biliary metal stent; Olympus Medical Systems, Tokyo, Japan) placement 7 months previously for unresectable pancreatic cancer. Endoscopic retrograde cholangiopancreatography (ERCP) was performed because recurrent biliary obstruction was suspected (► Fig. 1; ► Video 1). Cholangiography revealed defects suggestive of sludge (► Fig. 2). We therefore decided to remove the LC-CSEMS and replace it with a new one.

A snare forceps was used to hold the LC-CSEMS, which was moved toward the papilla by pushing, along with clockwise torsion of the endoscope while adjusting the axis of the bile duct [3]. The initial attempt to pull the LC-CSEMS out was effective; however, the snare subsequently could not be released, with the plan having originally been to release it from the LC-CSEMS and reposition it nearer the papilla (► Fig. 3a). Consequently, the LC-CSEMS was instead removed by withdrawing the endoscope. After stent removal, we inserted the endoscope to the papilla and found that a hemorrhage had been induced from the papilla during LC-CSEMS removal (► Fig. 3b). Although biliary cannulation was possible, the hemorrhage disturbed the endoscopic view, so re-insertion of the new LC-CSEMS was performed under fluoroscopic guidance. The LC-CSEMS suffers from minimal stent shortening because of the laser-cut structure; there-

► **Fig. 1** Endoscopic retrograde cholangiopancreatography findings showing: **a** the endoscopic appearance of the successfully placed laser-cut covered self-expandable metal stent (LC-CSEMS) for a patient with unresectable pancreatic cancer; **b** the fluoroscopic appearance of the successfully placed LC-CSEMS.

► **Video 1** Difficult removal of a laser-cut covered self-expandable metal stent (LC-CSEMS) causing hemorrhage from the papilla that was subsequently controlled by insertion of a replacement LC-CSEMS.

fore, it could be placed easily in an accurate position (► Fig. 4a) and the hemorrhage was finally stopped by compression from the LC-CSEMS (► Fig. 4b). Although this hemorrhage was induced by LC-CSEMS removal, it was easily stopped after the LC-CSEMS was replaced.

Endoscopy_UCTN_Code_CPL_1AK_2AD

Competing interests

The authors declare that they have no conflict of interest.

► **Fig. 2** Cholangiographic appearance with defects (pink arrows) suggestive of sludge.

Acknowledgments

We would like to thank Editage (www.editage.com) for English language editing.

The authors

Yuki Tanisaka **Masafumi Mizuide, Akashi Fujita** **Tomoya Ogawa, Hiromune Katsuda, Youichi Saito, Shomei Ryozaawa**
Department of Gastroenterology, Saitama Medical University International Medical Center, Hidaka, Saitama, Japan

Corresponding author

Yuki Tanisaka, MD, PhD

Department of Gastroenterology, Saitama Medical University International Medical Center, 1397-1, Yamane, Hidaka, Saitama 350-1298, Japan
tanisaka1205@gmail.com

References

- [1] Isayama H, Komatsu Y, Tsujino T et al. A prospective randomised study of “covered” versus “uncovered” diamond stents for the management of distal malignant biliary obstruction. *Gut* 2004; 53: 729–734
- [2] Kitano M, Yamashita Y, Tanaka K et al. Covered self-expandable metal stents with an anti-migration system improve patency duration without increased complications compared with uncovered stents for distal biliary obstruction caused by pancreatic carcinoma: a randomized multicenter trial. *Am J Gastroenterol* 2013; 108: 1713–1722

► **Fig. 3** Endoscopic views showing: **a** the snare forceps, which could not be released, holding the partially withdrawn laser-cut covered self-expandable metal stent; **b** hemorrhage that was induced during laser-cut covered self-expandable metal stent removal.

► **Fig. 4** Endoscopic retrograde cholangiopancreatography findings showing: **a** the fluoroscopic appearance of the successfully placed accurately positioned laser-cut covered self-expandable metal stent (LC-CSEMS); **b** the endoscopic appearance with hemostasis successfully achieved by compression from the LC-CSEMS.

- [3] Tanisaka Y, Ryozaawa S, Kobayashi M et al. Endoscopic removal of laser-cut covered self-expandable metallic biliary stents: A report of six cases. *Mol Clin Oncol* 2018; 8: 269–273
- [4] Marui S, Uza N, Yamazaki H et al. Utility of laser-cut covered self-expandable metal stents for unresectable malignant distal biliary obstruction: a single-center experience. *Endoscopy* 2020; 52: 664–668
- [5] Kin T, Ishii K, Okabe Y et al. Feasibility of biliary stenting to distal malignant biliary obstruction using a novel designed metal stent with duckbill-shaped anti-reflux valve. *Dig Endosc* 2021; 33: 648–655

Bibliography

Endoscopy 2022; 54: E378–E379
DOI 10.1055/a-1541-7205
ISSN 0013-726X
published online 9.8.2021
© 2021. Thieme. All rights reserved.
Georg Thieme Verlag KG, Rüdigerstraße 14, 70469 Stuttgart, Germany