Endoscopic treatment of transcolonic misplacement of a percutaneous endoscopic gastrostomy tube

Fig. 1 Percutaneous endoscopic gastrostomy (PEG) tube traversing the sigmoid at 40 cm from the anus.

Fig. 2 Abdominal transillumination at the site of the PEG entry.

Fig. 3 Computed tomography scan of the abdomen confirming the PEG tube running through the colon, with the internal bumper correctly lying within the stomach.

Fig. 4 Endoscopic closure of the colonic perforations with clips.

An 80-year-old man, who had received a percutaneous endoscopic gastrostomy (PEG) tube 5 months earlier, was referred to our institution for further evaluation of bloody diarrhea. Physical examination revealed a reduced level of consciousness and left-sided abdominal tenderness. Endoscopy demonstrated inflamed, edematous and partly necrotic colonic mucosa consistent with ischemic colitis. To our surprise, a PEG tube traversing the sigmoid at 40 cm from the anus was seen (▶ **Fig. 1**) with abdominal transillumination at the site of PEG entry (▶ **Fig. 2**). A computed tomography (CT) scan of the abdomen confirmed the PEG tube running through the colon with the internal bumper correctly lying within the stomach (▶ **Fig. 3**). In addition, thickening of the left-sided colonic wall was demonstrated. Three days later and after initiation of antibiotic therapy,

colonoscopy and gastroscopy were performed to remove the misplaced PEG tube (▶ **Video 1**). The PEG tube was gastroscopically retrieved with a snare. The removal was simultaneously observed with the colonoscope, and endoscopic clips were immediately placed to close the colonic perforations (▶ **Fig. 4**). Clips were also applied to close the gastric perforation from within the stomach. Leak testing

Video 1

Removal of the misplaced PEG tube under colonoscopic observation and endoscopic closure of the gastrocolonic and colocutaneous fistulas with clips.

with air and methylene blue injections through the colonoscope confirmed complete closures of the gastrocolic and colocutaneous fistulas, respectively. Finally, a new PEG tube was introduced using the pull technique following positive transillumination. The patient recovered rapidly and could be discharged on hospital day 7. Gastrocolonic fistula after PEG tube placement is an uncommon but serious complication. There are several case reports on the closure of such fistulas by surgery, fibrin glue [1] or septal repair implants [2]. In the present case, we could successfully close the fistula by placement of multiple clips immediately after retrieval of the PEG tube. Of interest here is that

the occurrence of ischemic colitis led to the detection of the misplaced PEG tube. Hence, one might speculate that this condition was related to the tight fixation of the sigma in the upper abdomen caused by the transcolonic PEG misplacement.

Endoscopy_UCTN_Code_CPL_1AH_2AI

**S. von Delius¹, W. Huber¹,
T. Dienemann¹, J. Gaa², M. Frandah³,
R. M. Schmid¹, A. Meining¹**

¹ 2nd Medical Department, Klinikum rechts der Isar der Technischen Universität München, Munich, Germany

- ² Department of Radiology, Klinikum rechts der Isar der Technischen Universität München, Munich, Germany
³ Department of Medicine, Al-Fatah University for Medical Sciences, Tripoli, Libya

References

- 1 González-Ojeda A, Avalos-González J, Muciño-Hernández MI et al. Fibrin glue as adjuvant treatment for gastrocutaneous fistula after gastrostomy tube removal. *Endoscopy* 2004; 36: 337–341
- 2 Melmed GY, Kar S, Geft I, Lo SK. A new method for endoscopic closure of gastrocolonic fistula: novel application of a cardiac septal defect closure device (with video). *Gastrointest Endosc* 2009; 70: 542–545

Bibliography

DOI 10.1055/s-0029-1243864

Endoscopy 2010; 42: E50–E51

© Georg Thieme Verlag KG Stuttgart · New York ·
ISSN 0013-726X

Corresponding author

A. Meining, MD

Technische Universität München

Klinikum rechts der Isar

2nd Medical Department

Ismaninger Str. 22

81675 Munich

Germany

Fax: +49-89-41404905

alexander.meining@lrz.tum.de