Endoscopic removal of the Padlock-G clip

In recent years various new endoscopic closure techniques and devices have been developed [1]. Research has mainly been driven by the need for a secure and reliable closure for natural orifice transluminal endoscopic surgery (NOTES) [2]. At present, the clinical applicability of NOTES remains unclear, but research in this area has already immeasurably enriched our endoscopic armamentarium with regards to endoscopic closure of perforations [1,3,4].

Recently, two novel clipping devices have been developed, which are conceptually similar to endoscopic band ligation [3–5]. One of these devices is the Padlock-G clip (▶ **Fig. 1**; Aponos Medical, Kingston, New Hampshire, USA), the feasibility of which has been demonstrated recently [5]. The closure mechanism consists of a 16.5-mm nitinol clip delivered via an over-the-scope delivery pod. Herein we report a technique that we have developed for the safe removal of this clip after it has been deployed.

In a 34-kg female domestic pig under general anesthesia, an 18-mm gastric wall opening was created using a needle knife and a dilation balloon. The Padlock-G clip was deployed after approximating the gastrotomy borders with a specialized tissue approximation grasper (Ovesco Endoscopy AG, Tübingen, Germany), thus creating a full-thickness closure of the defect (▶ **Fig. 2a**, ▶ **Video 1**). Time to achieve endoscopic closure was 3 minutes.

For removal, a soft oval endoscopic snare (SD-210U-25, Olympus, Center Valley, Pennsylvania, USA) was used. By grasping two of the side bars, each anchoring pin of the clip can be pulled out of the tissue in a serial fashion and the clip can be removed with minimal tissue trauma (▶ **Fig. 2b–d**, ▶ **Video 2**). Removal was facilitated within 1 minute and without complications.

Fig. 1 Padlock-G clip closure device. **a** The clip; **b** the loaded clip on the applicator cap.

Fig. 2 Removal of the Padlock-G clip. **a** The Padlock-G clip closure of an 18-mm full-thickness gastric wall defect. **b–d** Padlock-G clip removal using a standard endoscopic snare. By grasping two of the side bars (**b**), each anchoring pin of the clip can be pulled out of the tissue (**c**) in a serial fashion and the clip is removed (**d**).

In conclusion, the novel Padlock-G clip seems to be a promising new device for endoscopic organ wall closure with the additional benefit of easy and swift endoscopic removal in cases of unsatisfactory or incomplete closure attempts.

Competing interests: None

Endoscopy_UCTN_Code_TTT_1AO_2AI

Video 1

Padlock closure of an 18-mm gastric wall defect.

Video 2

Technique of endoscopic Padlock removal.

**D. von Renteln¹, M. C. Vassiliou²,
R. I. Rothstein³**

¹ Department of Interdisciplinary Endoscopy, University Hospital Hamburg-Eppendorf, Hamburg, Germany

² Department of Surgery, McGill University Health Centre, Montreal, Quebec, Canada

³ Section of Gastroenterology and Hepatology, Dartmouth-Hitchcock Medical Center, Lebanon, New Hampshire, USA

References

- 1 Voermans RP, Worm AM, van Berge Henegouwen MI et al. In vitro comparison and evaluation of seven gastric closure modalities for natural orifice transluminal endoscopic surgery (NOTES). *Endoscopy* 2008; 40: 595–601
- 2 ASGE/SAGES Working Group on Natural Orifice Transluminal Endoscopic Surgery. White Paper October 2005. *Gastrointest Endosc* 2006; 63: 199–203
- 3 Renteln D von, Schmidt A, Vassiliou MC et al. Endoscopic closure of large colonic perforations using an over-the-scope clip: a randomized controlled porcine study. *Endoscopy* 2009; 41: 481–486
- 4 Renteln D von, Rudolph HU, Schmidt A et al. Endoscopic closure of duodenal perforations by using an over-the-scope clip: a randomized, controlled porcine study. *Gastrointest Endosc* 2010; 71: 131–138
- 5 Romanelli JR, Desilets DJ, Earle DB. Natural orifice transluminal endoscopic surgery gastrotomy closure in porcine explants with the Padlock-G clip using the Lock-It system. *Endoscopy* 2010; 42: 306–310

Bibliography

DOI 10.1055/s-0030-1255620

Endoscopy 2010; 42: E241–E242

© Georg Thieme Verlag KG Stuttgart · New York · ISSN 0013-726X

Corresponding author

D. von Renteln, MD

Department of Interdisciplinary Endoscopy
University Hospital Hamburg-Eppendorf

Martinistr. 52
20251 Hamburg

Germany

Fax: +49-7141-997463

renteln@gmx.net