Evaluation of 19-gauge endoscopic ultrasonography aspiration needles using various echoendoscopes

Authors

Takao Itoi¹, Fumihide Itokawa¹, Atsushi Sofuni¹, Toshio Kurihara¹, Takayoshi Tsuchiya¹, Kentaro Ishii¹, Shujiro Tsuji¹, Nobuhito Ikeuchi¹, Hiroshi Kawakami², Fuminori Moriyasu¹, Ichiro Yasuda³

Institutions

¹ Department of Gastroenterology and Hepatology, Tokyo Medical University, Tokyo, Japan

² Department of Gastroenterology, Hokkaido University Graduate School of Medicine, Sapporo, Japan

³ First Department of Internal Medicine, Gifu University Hospital, Gifu

Bibliography

DOI <http://dx.doi.org/10.1055/s-0033-1359212>
Endoscopy International Open 2013; 1: E24–E30
© Georg Thieme Verlag KG
Stuttgart · New York
E-ISSN 2196-9736

Corresponding author:

Takao Itoi, M.D., FASGE
Department of
Gastroenterology and
Hepatology, Tokyo Medical
University
6-7-1 Nishishinjuku
Shinjuku-ku, Tokyo 160-0023
Japan
ittoi@tokyo-med.ac.jp

Background and study aims: 19-gauge (19G) needles are used for EUS-guided tissue acquisition and interventions. The aim of the current study was to compare the functional characteristics of 19G EUS needles by means of using various echoendoscopes in a bench simulator.

Methods: The angles achieved for 19G needles (EchoTip: ET-19G, EchoTip Flex: ExF-19G, Expect: Ex-19G, and ProCore: PC-19G) and for the distal tip of the echoendoscope were evaluated for maximal angulation settings of the distal tip and the elevator. Also the resistance to advancement of the 19G needles at these endoscope settings was assessed. All evaluations were done with endoscopes in a straight and in a curved position.

Results: There was no large discrepancy for scope and needle angles among all echoendoscopes except for the slim Pentax scope (EG-3270UK). The ExF-19G and PC-19G needles showed better optimal angles in various conditions compared to standard 19G needles. In straight scope position,

out of the 20 conditions (5 endoscopes × 2 positions of the distal tip × 2 elevator positions) the numbers of optimal angulations achieved for the Ex-19G, ExF-19G, ET-19G, and PC-19G, were 11/20 (55%), 20/20 (100%), 14/20 (70%) and 18/20 (90%), respectively. However, regarding resistance, it was impossible to advance the needle with 14/20 settings (70%) for the Ex-19G, 3/20 (15%) for the ExF-19G, 10/20 (50%) for the ET-19G and 7/20 (35%) for the PC-19G. When the scopes were bent, with regard to the force needed to advance the needle, the numbers of optimal settings, for the Ex-19G, ExF-19G, ET-19G, and PC-19G, were 1 (5%), 13 (65%), 6 (30%) and 8 (40%), respectively. The mean maximum resistance to advancement was less for the ExF-19G than for the other needles in almost all scope and angle conditions ($p < 0.05$).

Conclusion: Although there was no difference between needles, the resistance to passage was least with the flexible 19-gauge needle (ExF-19G).

Introduction

Several 19-gauge needles are currently available for endoscopic ultrasonography-guided fine-needle aspiration (EUS-FNA) and interventions [1–18]. However, the large bore and inherent stiffness of these needles hinder their maneuverability for puncture. In addition, other factors such as scope position (bent or curved), echoendoscope tip angulation, and elevator use affect the maneuverability of the needle.

In a prior study from our group that evaluated one particular 19G FNA needle (EchoTip, Cook Endoscopy, Winston-Salem, N.C.), we found that the needle was suboptimal for performing transduodenal FNA of pancreatic head and uncinate lesions [19]. We observed that when the tip of the echoendoscope was angulated or when the elevator was used, there was increased resistance to passage of the 19G needle that precluded sam-

pling of lesions. Recently, several 19-gauge needles and small diameter echoendoscopes have been developed. In this study, we compared the characteristics of different 19-gauge EUS needles using various echoendoscopes in a bench simulator (see below) to provide standardized, reproducible, comparative performance data.

Materials and methods

Echoendoscopes

Five curved linear array echoendoscopes (GF-UCT260, GF-UCT240, and GF-UC240P: Olympus Medical Systems, Tokyo, Japan; EG-3870UTK and EG-3270UK: Pentax, Tokyo, Japan), were examined using the bench simulator. The specifications of these echoendoscopes are described in

• **Table 1.**

License terms

	Olympus		Pentax	
	GF-UCT260 / GF-UCT240	GF-UC240P	EG-3870UTK	EG-3270UK
Outer diameter (mm)				
Distal end	14.6	14.2	14.3	12
Insertion tube	12.6	11.8	12.8	10.8
Bending section (degree)				
Up/down	130/90	130/90	130/130	130/130
Right/left	90/90	90/90	120/120	120/120
Working length (mm)	1,250	1,250	1,250	1,250
Working channel diameter (mm)	3.7	2.8	3.8	2.8
Elevator angle (degree)	90	90	90	90

Table 1 Specifications of echoendoscopes.

Fig. 1 Bench simulator: **a** straight therapeutic scope position; **b** curved scope position.

19G needles

The characteristics of commercially available 19-gauge EUS needles were evaluated by bench simulation for both extent of possible angulation and for resistance to passage under various conditions. Four 19-gauge EUS needles were evaluated: 1) Expect™ standard type (Ex-19G) (Boston Scientific Japan, Tokyo, Japan), 2) Expect™ flexible type (ExF-19G) (Boston Scientific Japan), 3) EchoTip® ULTRA (ET-19G) (Cook Endoscopy), 4) EchoTip® ProCore™ (PC-19G) (Cook Endoscopy).

Before the bench simulation, the diameter of each needle was accurately measured by a digital micrometer caliper (MDC-25, Mitutoyo, Kanagawa, Japan).

Bench simulation

The bench simulator was designed and constructed by one of the authors (T.I.), specifically for the comparative evaluation of different EUS-FNA needles as previously described [19]. The operating portion of the echoendoscope was fixed using a vise, and the insertion portion was fixed in straight (Fig 1a) and curved (Fig 1b) positions using a wooden board, vinyl chloride tubing with a diameter of 19 mm (ST-19, Toyoc Co., Ltd., Toyama, Japan), and stainless steel saddle grips.

We repeated all the tests with the endoscopes held in both straight and curved positions.

Range of angulation of echoendoscope distal tips and consequent angulation of needles

We measured the maximum achievable angulation for each possible combination of the five echoendoscopes and the four 19G needles. The maximum distal tip “up” and elevator “up” settings were used. Thus the “big knob” (up-down) and not the small knob (left-right) of the echoendoscope was used. This was set at 0° and at maximum in each echoendoscope and the elevator was set at 0° and 90° (maximum). When the tip of the needle was flexed, the angle between the echoendoscope and the sheath of the needle was measured.

The echoendoscope angle and the needle angle, were defined as angle A and angle B, respectively, as shown in Fig 2. Optimal angulation was defined as angle A and angle B being more than 90° and 110°, respectively.

All measurements were repeated with the scopes held in both the “straight” and the “curved” positions (see Fig 1).

Resistance of needles

The resistance to needle advancement through the endoscope was measured using an Autograph (AGS-H, Shimadzu Corporation, Kyoto, Japan), which has a highest tolerable force of 1,000 newtons (1,000 N) (Fig 3). The needles were advanced at a speed of 500 mm/min. The length of the needle insertion

was 30 mm. The resistance to advancement of the needle was recorded continuously during insertion. The data were analyzed by dedicated software (Factory Shikibu 2000, Shimazu Factory Manufacture).

In this study, the resistance of each FNA needle was measured more than 3 times to obtain reproducible results (mean 3.2 times, range 3 to 5 times) using each scope at each of the possible settings. We prepared and used five each of the four 19G needles for the examination.

As there were five scopes, each with two settings of the distal tip (up/down) and two settings of the elevator (up/down), 20 values were obtained for each needle. Again, all of these measurements were repeated with the scopes held in both the “straight” and the “curved” positions (see [Fig 1](#)).

We used our prior data for resistance to needle advancement (R_{needle}) [18] for grading the resistance: $R_{\text{needle}} < 8$ N, optimal setting; $R_{\text{needle}} > 8$ N, suboptimal; $R_{\text{needle}} > 11$ N, impossible to advance the needle.

Statistics

Statistical analyses were performed with StatMate III (ATMS Co. Ltd., Tokyo, Japan). Normally distributed data are presented as the mean $\pm$ standard deviation (SD). Kruskal-Wallis test as non-parametric statistic was used to compare between each needle resistances. A P value less than 0.05 was regarded as indicating a statistically significant difference.

Results

Actually achieved distal tip and needle angulations Type of echoendoscopes

Olympus The results for all the combinations of scopes and needles, with the scope held in a straight position, are shown in [Table 2](#). In the straight scope position, for both angle A and angle B, the new therapeutic Olympus scope (GF-UCT260) showed optimal angulation using various 19-gauge needles ([Table 2](#)). However, both angle A and angle B in two echoendoscopes (GF-UCT240 and GF-UCP240) using standard FNA needles (Ex-19G and ET-19G) were suboptimal. In the curved scope position, angle A and angle B were each optimal in only one of the 16 possible settings. In both the straight and the curved scope position, the GF-UCT260 allowed in 50% of possible settings better optimal angles compared to others ([Table 3](#)).

Pentax Regardless of scope positions, the therapeutic Pentax scope (EG-3870UTK) showed optimal angles compared to the slim echoendoscope (EG-3270UK) ([Table 3](#)).

The type of needles

Standard FNA needles (Ex-19G and ET-19G) In the straight scope position, using the Ex-19G showed optimal angles A and B in only two echoendoscopes the (GF-UCT260 and the EG-3870UTK). With the curved scope position, in almost all conditions, angles A and B were suboptimal angles. The results for the ET-19G were similar to those of the EX-19G.

PC-19G In the straight scope position, the PC-19G showed optimal angle A and B in all but one echoendoscope (EG-3270UK). Similarly, with the curved scope position, angle B was optimal in all but one echoendoscope (EG-3270UK) though angle A in three echoendoscopes (GF-UCT240, GF-UCP240 and EG-3270UK) was suboptimal.

ExF-19G When the scope was straight, the ExF-19G showed optimal Angle A and B in all but one echoendoscope (EG-3270UK). With the curved scope position, 5 out of the 10 values for angles A and B were optimal.

Resistance to advancement of 19-gauge needles The maximum resistance to advancement of the FNA needles is shown in [Table 4](#) for the straight scope position and in [Table 5](#) for the curved scope position. The results show high reproducibility with only small SDs ([Table 4](#) and [Table 5](#)). In all cases, the resistance to the passage of the FNA needles, R_{needle} , was greater with greater angulation of the distal tip and of the elevator. Increased resistance was observed irrespective of the type of echoendoscope.

Straight scope position

With a straight scope, the numbers of optimal settings of the distal tip and the elevator in the Ex-19G, ExF-19G, ET-19G, and PC-19G, were 11/20 (55%), 20/20 (100%), 14/20 (70%) and 18/20 (90%), respectively ([Table 4](#)). These were the settings where the resistance to advancement of the needle was ≤ 8 N. The mean maximum R_{needle} was smaller for the ExF-19G than for others in all scope and angle conditions ($p < 0.05$).

Curved scope position

With the scope in a curved position, needle advancement was graded as “impossible” in 14/20 settings (70%) for the Ex-19G, 3/20 (15%) for the ExF-19G, 10/20 (50%) for the ET-19G and 7/20 (35%) for the PC-19G ([Table 5](#)). On the other hand, the numbers of optimal settings for the Ex-19G, ExF-19G, ET-19G, and PC-19G, were 1 (5%), 13 (65%), 6 (30%) and 8 (40%), respectively. In almost all conditions, the mean maximum R_{needle} for the ExF-19G was smaller than that of others ($p < 0.05$).

Discussion

Interventional EUS procedures frequently require the use of a 19-gauge needle in conjunction with a 0.035-inch guidewire. However, most 19-gauge needles, given their large caliber, encounter resistance when the echoendoscope tip is angulated resulting in poor maneuverability. Our experimental study [19] revealed that the distal tip up and elevator-up positions lead to higher resistance for the needle, resulting in difficulty in puncture by the needle and contributing to shearing of the guidewire. To overcome this limitation, recently, newer 19-gauge needles and a slim echoendoscope (the EG-3270UK) have been developed. Until now, although there have been several experimental and clinical comparative studies between 19-gauge needle and others (Trucut, 22- and 25-gauge needles) [19–23], there has been no comparative study for various 19-gauge needles including newer ones in combination with various old and newer echoendoscopes.

Angles A and B may represent, respectively, the ability to delineate and to puncture targeted lesions. Smaller angles may mean less needle maneuverability, thereby limiting the various interventions. In the present study, although initially we postulated that there was no difference between the angles A and B among the standard echoendoscopes equipped with a 19G needle, the standard Pentax scope (EG-3870UTK) showed optimal angulation with a variety of 19G needles compared with standard Olympus echoendoscopes. However, we should consider actual clinical situations because the echoendoscopic delineation of tar-

Fig. 2 Measurement of angulations of distal tip of echoendoscope and needle: A, distal tip angulation; B, needle angulation.

Fig. 3 Autograph equipment for examination of resistance to needle insertion.

geted lesions depends on both the flexibility of the scope and the direction of the ultrasound probe. The direction of the ultrasound probe differs between the Pentax and Olympus echoendoscopes, which have a side and an oblique direction, respectively, for the ultrasound beam direction. Therefore, the actual differentiation between the standard Pentax and Olympus scopes may be unclear because we did not image any real lesions by means of EUS with a 19G needle. The diagnostic slim Pentax echoendoscope (EG-3270UK) showed comparatively lower scope angulations than others. These data suggest that the standard large-diameter EUS scope has a robust wire that enables the large tip of the endoscope to be safely bent, while the diagnostic slim echoendoscope does not have a such a robust wire. That may lead to lower angulations when the 19-gauge needle is used. Since the needle angulation (angle B) is the most important parameter for various interventional EUS procedures, the Pentax slim echoendoscope in combination with a 19-gauge needle seems to be suboptimal for interventional EUS.

In terms of selection of a 19-gauge needle, from the aspect of the scope-needle angles allowed by various echoendoscopes, our data suggested that among the various 19-gauge needles, the Expect™ flexible type needle (Boston Scientific) and/or the ProCore™ needle (Cook Endoscopy) appear to be optimal for interventional EUS including EUS-FNA, because of the optimal angulation in various scope settings. However, the two needles are quite different in concept: the ProCore™ needle is basically used for dedicated EUS-FNA procedures and the Expect™ flexible needle for both FNA and interventions. In terms of cytological/histological performance, we suggest that there may be differences between the classic 19-G needle and the ProCore™ 19-G/ Expect™ 19-G flexible type needle in special settings, e.g. the trans-duodenal approach to lesion puncture. In addition, they seem to be optimal for EUS-FNA when core tissues, obtained using a 19-G needle are preferable for accurate diagnosis in special settings such as malignant lymphoma and autoimmune pancreatitis, although current 22-G and/or 25-G needles are sufficient for EUS-FNA in patients with common diseases such as pancreatic cancers.

Even when lesions can be well delineated, the puncture is not always successful because of lack of sharpness of the needle and the resistance to its advancement. Resistance to the needle is caused by various factors, e.g. torque, kinking of working channel due to scope position, and use of distal tip/elevator angulation. These limited the performance of the needle, resulting in failure of puncture during EUS-FNA. Our previous study evaluating resistance to the needle revealed that resistance to needle advancement was moderate at a resistance of 5 N or more and high at 8 N or more, and that it was impossible to advance the needle when the resistance was 11 N or more [19]. In the short scope position, the transgastric approach, unless the maximum up-angulation was applied in the distal tip and the elevator at the same time, needle punctures was possible using any 19-gauge needle. The results of the present study were similar to our previous data [19]. In the straight scope position, the new flexible 19-gauge needle encountered comparatively less resistance than other needles and all needles encountered lower resistance (<8 N). When the scope was curved, most of the 19G needles met resistance greater than 10 N. In particular, in the case of both distal-tip-up and elevator-up angulation, resistances to the needle were more than 11 N, meaning that puncture was impossible. Also surprisingly, the R_{needle} for the new flexible 19-gauge needle was lower than that of others. Our results suggest that for needle

Table 2 Angles of echoendoscope and needle using the various 19-gauge EUS needles in the straight scope position.

	Angle of echoendoscope (angle A) (degree)				Angle of needle (angle B) (degree)			
	Ex-19G	ExF-19G	ET-19G	PC-19G	Ex-19G	ExF-19G	ET-19G	PC-19G
Olympus								
GF-UCT260	93	96	90	94	113	115	114	116
GF-UCT240	77	90	88	92	96	115	108	115
GF-UCP240	85	92	88	90	105	115	105	115
Pentax								
EG-3870UTK	90	100	100	110	113	125	135	136
EG-3270UK	55	70	72	80	95	95	104	104

Abbreviations: ET, EchoTip; EUS, endoscopic ultrasonography; EX, Expect; ExF, Expect Flex; G, gauge; PC, ProCore.

Table 3 Angles of echoendoscope and needle using the various 19-gauge EUS needles in the curved scope position.

	Angle of echoendoscope (angle A) (degree)				Angle of needle (angle B) (degree)			
	Ex-19G	ExF-19G	ET-19G	PC-19G	Ex-19G	ExF-19G	ET-19G	PC-19G
Olympus								
GF-UCT260	92	95	82	90	108	112	100	115
GF-UCT240	78	90	84	87	104	104	106	115
GF-UCP240	80	80	80	88	105	104	106	111
Pentax								
EG-3870UTK	85	95	88	92	105	124	125	130
EG-3270UK	55	65	68	70	75	90	94	91

Abbreviations: ET, EchoTip; EUS, endoscopic ultrasonography; EX, Expect; ExF, Expect Flex; G, gauge; PC, ProCore.

Table 4 The maximum resistance to advancement of needle in the straight scope position (in newtons, N, mean $\pm$ SD).

Echoendoscope	Distal tip angulation	Elevator angulation	Ex-19G	ExF-19G	ET-19G	PC-19G
Olympus						
GF-UCT260	0°	0°	5.03 $\pm$ 0.53	2.01 $\pm$ 0.21	3.04 $\pm$ 0.05	3.35 $\pm$ 0.09
		Max	10.73 $\pm$ 0.34	4.10 $\pm$ 0.20	6.60 $\pm$ 0.43	7.03 $\pm$ 0.13
	Max	0°	7.78 $\pm$ 0.27	4.75 $\pm$ 0.53	5.04 $\pm$ 0.14	6.78 $\pm$ 0.15
		Max	11.55 $\pm$ 0.52	5.13 $\pm$ 0.13	6.12 $\pm$ 0.10	7.59 $\pm$ 0.22
GF-UCT240	0°	0°	4.68 $\pm$ 0.20	2.75 $\pm$ 0.10	4.48 $\pm$ 0.16	2.81 $\pm$ 0.17
		Max	8.60 $\pm$ 0.43	4.34 $\pm$ 0.19	8.28 $\pm$ 0.40	6.23 $\pm$ 0.32
	Max	0°	7.00 $\pm$ 0.21	4.98 $\pm$ 0.45	7.48 $\pm$ 0.21	5.05 $\pm$ 0.10
		Max	9.47 $\pm$ 0.12	5.61 $\pm$ 0.36	9.09 $\pm$ 0.19	6.93 $\pm$ 0.08
GF-UCP240	0°	0°	7.82 $\pm$ 0.25	1.61 $\pm$ 0.09	4.38 $\pm$ 0.18	3.14 $\pm$ 0.43
		Max	9.30 $\pm$ 0.37	3.78 $\pm$ 0.21	6.93 $\pm$ 0.03	5.58 $\pm$ 0.08
	Max	0°	13.63 $\pm$ 0.40	3.77 $\pm$ 0.08	7.58 $\pm$ 0.52	5.40 $\pm$ 0.09
		Max	14.58 $\pm$ 0.20	5.60 $\pm$ 0.87	8.58 $\pm$ 0.10	6.87 $\pm$ 0.10
Pentax						
EG-3870UTK	0°	0°	4.23 $\pm$ 0.25	1.57 $\pm$ 0.08	3.12 $\pm$ 0.06	2.83 $\pm$ 0.16
		Max	6.02 $\pm$ 0.39	2.98 $\pm$ 0.03	6.05 $\pm$ 0.05	4.77 $\pm$ 0.08
	Max	0°	7.98 $\pm$ 0.30	3.53 $\pm$ 0.33	6.83 $\pm$ 0.13	4.28 $\pm$ 0.53
		Max	9.30 $\pm$ 0.99	4.05 $\pm$ 0.23	8.55 $\pm$ 0.23	7.03 $\pm$ 0.10
EG-3270UK	0°	0°	3.63 $\pm$ 0.21	1.57 $\pm$ 0.03	4.10 $\pm$ 0.15	2.82 $\pm$ 0.13
		Max	5.00 $\pm$ 0.17	7.30 $\pm$ 0.64	10.90 $\pm$ 0.20	8.42 $\pm$ 0.23
	Max	0°	7.95 $\pm$ 0.30	3.68 $\pm$ 0.08	7.33 $\pm$ 0.08	5.23 $\pm$ 0.13
		Max	8.88 $\pm$ 0.33	7.42 $\pm$ 1.01	10.93 $\pm$ 0.17	9.32 $\pm$ 0.13

Abbreviations: ET, EchoTip; Ex, Expect; ExF, Expect Flex; Max, maximum angulation; PC, ProCore.

Table 5 The maximum resistance to advancement of needle in the curved scope position (in newtons N, mean $\pm$ SD)

Echoendoscope	Distal tip angulation	Elevator angulation	Ex-19G	ExF-19G	ET-19G	PC-19G
Olympus						
GF-UCT260	0°	0°	10.29 ± 0.32	2.51 ± 0.17	5.23 ± 0.15	3.83 ± 0.22
		Max	13.52 ± 0.49	7.84 ± 0.52	12.78 ± 0.51	11.75 ± 0.14
	Max	0°	10.61 ± 0.74	5.29 ± 0.09	7.39 ± 0.15	7.29 ± 0.23
		Max	14.53 ± 0.77	9.16 ± 0.49	12.80 ± 0.34	11.91 ± 0.63
GF-UCT240	0°	0°	10.13 ± 0.62	3.05 ± 0.17	6.91 ± 0.54	4.61 ± 0.34
		Max	12.37 ± 0.38	10.55 ± 0.34	16.35 ± 0.76	12.08 ± 0.97
	Max	0°	13.80 ± 0.22	7.47 ± 0.10	10.65 ± 0.66	8.33 ± 0.34
		Max	15.23 ± 0.33	11.42 ± 0.08	17.35 ± 0.44	14.43 ± 0.86
GF-UCP240	0°	0°	8.87 ± 0.32	3.39 ± 0.20	7.66 ± 0.23	4.53 ± 0.18
		Max	12.05 ± 0.40	9.25 ± 0.15	14.50 ± 0.83	10.10 ± 0.15
	Max	0°	12.45 ± 0.26	6.23 ± 0.06	10.22 ± 0.20	7.60 ± 0.21
		Max	14.70 ± 0.25	12.15 ± 0.72	15.11 ± 0.71	11.67 ± 0.24
Pentax						
EG-3870UTK	0°	0°	10.77 ± 0.12	2.03 ± 0.08	5.64 ± 0.07	3.15 ± 0.10
		Max	12.62 ± 0.15	5.43 ± 0.10	12.30 ± 0.30	9.33 ± 0.53
	Max	0°	14.95 ± 0.48	4.55 ± 0.09	9.18 ± 0.08	7.10 ± 0.05
		Max	16.00 ± 0.35	5.88 ± 0.26	11.35 ± 0.20	9.95 ± 0.36
EG-3270UK	0°	0°	7.72 ± 0.53	2.47 ± 0.08	4.47 ± 0.23	3.33 ± 0.10
		Max	11.27 ± 0.31	10.11 ± 0.55	15.80 ± 0.22	14.78 ± 0.28
	Max	0°	11.65 ± 0.54	6.65 ± 0.17	9.70 ± 0.23	8.05 ± 0.22
		Max	13.00 ± 0.18	12.08 ± 0.40	17.77 ± 0.63	13.12 ± 0.51

Abbreviations: ET, EchoTip; EX, Expect; ExF, Expect Flex; Max, maximum angulation; PC, ProCore.

puncture in difficult situations, such as trans-duodenal approach, the flexible 19G needle may be most appropriate. However, even when the flexible 19-gauge needle is used, the tight bending of distal tip-up and elevator-up angulation hinders advancement of the needle and at times, puncture by the needle is impossible, in particular when diagnostic thinner echoendoscopes (GF-UCP240 and EG-3270UK) are used. We speculate that in such thinner scopes, the needle compliance decreases because of kinking of the smaller working channel by distal tip angulation, or because of strong friction when the needle is pinned by the elevator mechanism.

The up-to-date findings of our study may be important because they are an important step in knowing which 19-gauge needles and endoscopes are better for EUS-related procedures. Our results suggest that both the Expect™ flexible type needle and the ProCore™ needle seem to be optimal for EUS-FNA used with either therapeutic Olympus (GF-UCT260 and GF-UCT240) or Pentax scopes (EG-3870UTK). However, for interventional EUS that requires the use of a guidewire, the Expect™ flexible needle appears to be optimal, in particular for the trans-duodenal approach.

The current 19-gauge needles, however, are still not the best for EUS-related procedures. For instance, the 19G ProCore™ needle and the Expect™ 19G flexible needle are still not as flexible compared to current 22- and 25-gauge needles. Thus, we still need of better technology to overcome this limitation.

There are several limitations in this study. Since various factors, such as echoendoscope position, the patient's pathology, anatomical differences, and torque of the scope cannot be reproduced in an experimental model, the conclusions may not hold true for all clinical scenarios.

In conclusion, our experimental data suggests that both the Expect™ flexible type needle and the ProCore™ needle seems to be most suitable for EUS-FNA procedures. The Expect™ flexible

type needle appears to be particularly suitable for the trans-duodenal approach and for interventional EUS procedures.

Competing interests: None.

Acknowledgement

The authors are indebted to Professor J. Patrick Barron of the International Medical Communications Center of Tokyo Medical University for his review of this manuscript.

References

- 1 Iwashita T, Yasuda I, Doi S et al. Use of samples from endoscopic ultrasound-guided 19-gauge fine-needle aspiration in diagnosis of autoimmune pancreatitis. *Clin Gastroenterol Hepatol* 2012; 10: 316–322
- 2 Varadarajulu S, Bang JY, Hebert-Magee S. Assessment of the technical performance of the flexible 19-gauge EUS-FNA needle. *Gastrointest Endosc* 2012; 76: 336–343
- 3 Varadarajulu S, Lopes TL, Wilcox CM et al. EUS versus surgical cyst-gastrostomy for management of pancreatic pseudocysts. *Gastrointest Endosc* 2008; 68: 649–655
- 4 Itoi T, Binmoeller KF, Shah J et al. Clinical evaluation of a novel lumen-apposing metal stent for endosonography-guided pancreatic pseudocyst and gallbladder drainage (with video). *Gastrointest Endosc* 2012; 75: 870–876
- 5 Park DH, Lee SS, Moon SH et al. Endoscopic ultrasound-guided versus conventional transmural drainage for pancreatic pseudocysts: a prospective randomized trial. *Endoscopy* 2009; 41: 842–848
- 6 Seewald S, Ang TL, Richter H et al. Long-term results after endoscopic drainage and necrosectomy of symptomatic pancreatic fluid collections. *Dig Endosc* 2012; 24: 36–41
- 7 Giovannini M, Dotti M, Bories E et al. Hepaticogastrostomy by echoendoscopy as a palliative treatment in a patient with metastatic biliary obstruction. *Endoscopy* 2003; 35: 1076–1078
- 8 Horaguchi J, Fujita N, Noda Y et al. Endosonography-guided biliary drainage in cases with difficult transpapillary endoscopic biliary drainage. *Dig Endosc* 2009; 21: 239–244

- 9 Itoi T, Isayama H, Sofuni A et al. Stent selection and tips on placement technique of EUS-guided biliary drainage: transduodenal and trans-gastric stenting. *J Hepatobiliary Pancreat Sci* 2011; 18: 664–672
- 10 Artifon EL, Pinhata Otoch J et al. Endoscopy Ultra- Sonography -guided biliary drainage in the surgical -endoscopy era. *Rev Gastroenterol Peru* 2001; 31: 365–375
- 11 Park DH, Koo JE, Oh J et al. EUS-guided biliary drainage with one-step placement of a fully covered metal stent for malignant biliary obstruction: A prospective feasibility study. *Am J Gastroenterol* 2009; 104: 2168–2174
- 12 Hara K, Yamao K, Niwa Y et al. Prospective clinical study of EUS-guided choledochoduodenostomy for malignant lower biliary tract obstruction. *Am J Gastroenterol* 2011; 106: 1239–1245
- 13 Kahaleh M. Therapeutic and advanced endoscopic retrograde cholangiopancreatography. *Gastrointest Endosc Clin N Am*. 2012 Jul22: xv
- 14 Baron TH, Topazian MD. Endoscopic transduodenal drainage of the gallbladder: implications for endoluminal treatment of gallbladder disease. *Gastrointest Endosc* 2007; 65: 735–737
- 15 Eum J, Park doHyun, Ryu CH et al. EUS-guided biliary drainage with a fully covered metal stent as a novel route for natural orifice transluminal endoscopic biliary interventions: a pilot study (with videos). *Gastrointest Endosc* 2011; 72: 1279–1284
- 16 Itoi T, Itokawa F, Kurihara T. Endoscopic ultrasonography-guided gallbladder drainage: actual technical presentations and review of the literature (with videos). *J Hepatobiliary Pancreat Sci* 2011; 18: 282–286
- 17 Giovannini M. Endoscopic ultrasonography-guided pancreatic drainage. *Gastrointest Endosc Clin N Am* 2012; 22: 221–230
- 18 Itoi T, Kikuyama M, Ishii K et al. EUS-guided rendezvous with single-balloon enteroscopy for treatment of stenotic pancreaticojejunal anastomosis in post-Whipple patients (with video). *Gastrointest Endosc* 2011; 73: 398–401
- 19 Itoi T, Itokawa F, Kurihara T et al. Experimental endoscopy: objective evaluation of EUS needles. *Gastrointest Endosc* 2009; 69: 509–516
- 20 Varadarajulu S, Fraig M, Schmulewitz N et al. Comparison of EUS-guided 19-gauge trucut needle biopsy with EUS-guided fine-needle aspiration. *Endoscopy* 2004; 36: 397–401
- 21 Itoi T, Itokawa H, Sofuni A et al. Puncture of solid pancreatic tumors guided by endoscopic ultrasonography: A pilot study series comparing trucut and 19-gauge and 22-gauge aspiration needle. *Endoscopy* 2005; 37: 362–366
- 22 Song TJ, Kim J, Lee SS et al. The prospective randomized, controlled trial of endoscopic ultrasound-guided fine-needle aspiration using 22G and 19G aspiration needles for solid pancreatic or peripancreatic masses. *Am J Gastroenterol* 2010; 105: 1739–1745
- 23 Sakamoto H, Kitano M, Komaki T et al. Prospective comparative study of the EUS guided 25-gauge FNA needle with the 19-gauge Trucut needle and 22-gauge FNA needle in patients with solid pancreatic masses. *J Gastroenterol Hepatol* 2009; 24: 384–390