Candy-like sign during endoscopic ultrasound-guided choledochoduodenostomy as an indication of the long distance between the bile duct and duodenal wall

Fig. 1 Radiographic image showing the candy-like sign of a partially covered self-expandable metallic stent resulting from a large gap between the common bile duct and the duodenal wall during endoscopic ultrasound-guided choledochoduodenostomy.

Fig. 2 Endoscopic image showing migration of the distal end of the partially covered self-expandable metallic stent during endoscopic ultrasound-guided choledochoduodenostomy.

Fig. 3 Radiograph showing balloon dilation of the first partially covered self-expandable metallic stent.

Endoscopic ultrasound-guided choledochoduodenostomy (EUS-CDS) using a covered self-expandable metallic stent (SEMS) is an established alternative drainage technique for patients in whom endoscopic retrograde cholangiopancreatography has failed [1,2]. This report describes

the case of a patient who underwent successful EUS-CDS with a partially covered SEMS placed far from the duodenal wall. A 66-year-old man with locally advanced pancreatic head cancer was admitted to our hospital. He had undergone percutaneous transhepatic biliary drainage

(PTBD) at another hospital 1 week earlier because of failed selective bile duct cannulation. He experienced right flank pain after PTBD. As the patient wanted the PTBD tube removed, we opted to perform EUS-CDS rather than antegrade stenting. After puncture of the common bile duct from the first part of the duodenum using a 19-gauge needle (Echo Tip Ultra; Cook Japan, Tokyo, Japan), a 0.025-inch guidewire (VisiGlide; Olympus Medical Systems Corp., Tokyo, Japan) was inserted into the intrahepatic bile duct. Fistula dilation was then performed using a 6-Fr wire-guided diathermic dilator (Cysto-Gastro-Set; Endo-Flex GmbH, Voerde, Germany) with a blended cut mode. Insertion of a partially covered SEMS (WallFlex, 10×60 mm; Boston Scientific Japan, Tokyo, Japan) was then attempted through the fistula. The EUS-CDS procedure was performed quickly; however, the SEMS revealed a candy-like sign in the form of a large gap between the bile duct and duodenum, and this sign warrants caution as it indicates distal migration and bile leakage (▶ Fig. 1 and ▶ Fig. 2; ▶ Video 1).

We therefore attempted additional stenting using a fully covered SEMS (Bonastent, 10×60 mm; Standard Sci Tech, Seoul, Korea), but this stent could not be passed through the first partially covered stent. Balloon dilation (Hurricane RX Biliary Bal-

Fig. 4 Radiograph showing the second fully covered self-expandable metallic stent (SEMS) placed across the first partially covered SEMS.

Fig. 5 Endoscopic image showing the second fully covered self-expandable metallic stent (SEMS) advanced over a guidewire through the first partially covered SEMS.

loon Dilation Catheter; Boston Scientific Japan) of the first partially covered SEMS was performed (● Fig. 3; ● Video 1). Finally, a second fully covered SEMS was placed across the fistula through the first partially covered stent without any complication (● Fig. 4 and ● Fig. 5; ● Video 1). We were successful in saving this patient using placement of an additional fully covered SEMS. Caution should be taken in the event of the rare and dangerous candy-like sign, which indicates a long distance between the bile duct and duodenum, during EUS-CDS with covered SEMS placement.

Video 1

Endoscopic rescue technique for gap formation of the partially covered self-expandable metallic stent (SEMS), which was dilated using a balloon catheter followed by placement of a fully covered SEMS.

Endoscopy_UCTN_Code_CPL_1AL_2AD

Competing interests: None

Hiroshi Kawakami, Masaki Kuwatani, Kazumichi Kawakubo, Taiki Kudo, Yoko Abe, Kimitoshi Kubo, Yoshimasa Kubota, Naoya Sakamoto

Department of Gastroenterology and Hepatology, Hokkaido University Graduate School of Medicine, Sapporo, Japan

References

- 1 Itoi T, Isayama H, Sofuni A et al. Stent selection and tips on placement technique of EUS-guided biliary drainage: transduodenal and transgastric stenting. *J Hepatobiliary Pancreat Sci* 2011; 18: 664–672
- 2 Kawakubo K, Isayama H, Kato H et al. A multicenter retrospective study of endoscopic ultrasound-guided biliary drainage (EUS-BD) for malignant biliary obstruction in Japan. *J Hepatobiliary Pancreat Sci* 2014; 21: 328–334

Bibliography

DOI <http://dx.doi.org/10.1055/s-0034-1377390>
Endoscopy 2014; 46: E406–E407
© Georg Thieme Verlag KG
Stuttgart · New York
ISSN 0013-726X

Corresponding author

Hiroshi Kawakami, MD, PhD
Department of Gastroenterology and Hepatology
Hokkaido University Graduate School of Medicine
Kita 15, Nishi 7
Kita-ku
Sapporo 060-8638
Japan
Fax: +81-11-7067867
hiropn@med.hokudai.ac.jp