Evaluation of Blood Loss and Implant Alignment after Total Knee Arthroplasty with Inertial Based Extramedullary Femoral Cutting Guide

Tommaso Bonanzinga^{1,2} Piergiuseppe Tanzi³ Maria Pia Neri³ Francesco Iacono^{1,2} Claudio Mazzola⁴
Alberto Belluati⁵ Andrea Colombelli⁵ Stefano Zaffagnini³ Maurilio Marcacci^{1,2}

¹ Department of Biomedical Sciences, Humanitas University, Milan, Italy

² Humanitas Clinical and Research Center, Milan, Italy

³ Rizzoli Orthopedic Institute, Bologna, Italy

⁴ Hospital Galliera, Genoa, Italy

⁵ Complex Unit of Orthopedic Hospital of Ravenna, Ravenna, Italy

Address for correspondence Tommaso Bonanzinga, MD, Humanitas Clinical and Research Center, Building 8, Via Manzoni 113, 20089 Rozzano, Milan, Italy (e-mail: t.bonanzinga@gmail.com).

Joints 2018;6:161–166.

Abstract

Purpose The purpose of this study was to compare, in terms of blood loss and implant alignment, a new generation of smart extramedullary (EM) cutting guides with the conventional intramedullary (IM) guide for total knee arthroplasty (TKA). The hypothesis was that the EM system would result in less blood loss and fewer days of hospitalization, while ensuring equal or higher accuracy in the alignment of the femoral implant.

Methods Thirty-six patients were enrolled for the present study: 18 patients underwent TKA using the EM guide and for the other 18 patients the IM guide was used. Preoperative and postoperative X-rays were acquired. The preoperative and postoperative hemoglobin values were compared. Lastly, length of hospital stay was recorded.

Results The hemoglobin difference was significantly lower in the EM group. The alignment of the prosthetic femoral implant in the two groups was comparable but all patients in the EM cohort had a final alignment within 3 degrees of range on the frontal view, while 22% of the patients in the IM cohort had a final alignment exceeding 3 degrees of range. The mean hospitalization duration showed no significant difference between groups, though EM group showed 1.6 days less compared with IM group.

Conclusions The inertial-based EM guide system could be useful for decreasing blood loss compared with conventional guide. This device tended to have better results even in terms of accuracy, but this difference was not significant.

Level of Evidence This is a level II, prospective cohort study.

Keywords

- total knee arthroplasty
- extramedullary guide
- blood loss
- alignment
- cutting guides

Introduction

Proper alignment of femoral, tibial, and patellar components is critical for implant survival of total knee arthroplasty (TKA). The desired alignment of the femoral component is traditionally obtained with the use of an intramedullary (IM) guide, but recent reports showed that IM guides achieve correct femoral component alignment, within less than

2-degree error, in less than 66% of knees.^{1,2} Moreover, the use of IM guides implies risks for the patients, such as perioperative blood loss and the requirement for blood transfusion after joint replacement. Kim et al³ showed that the invasiveness of IM guide leads to thromboembolic complications in 4% of cases with subsequent perioperative morbidity and neurological dysfunctions. Kalairajah et al⁴ reported a significant reduction in blood drainage and

received
November 18, 2017
accepted after revision
August 26, 2018
published online
October 18, 2018

DOI <https://doi.org/10.1055/s-0038-1673404>.
ISSN 2282-4324.

Copyright © 2018 Georg Thieme Verlag
KG Stuttgart · New York

License terms

hemoglobin loss with the use of computer-assisted surgery (CAS) in comparison with the conventional techniques. This allows reduced risks of transfusion and financial saving. Kandel et al⁵ reported that extramedullary (EM) femoral alignment instrumentation reduces blood loss after cementless TKA compared with IM device. In recent years, EM accelerometer-based devices were developed to improve accuracy of implant positioning⁶⁻⁹ and to reduce patient morbidity. These devices have inertial electronic components attached to the femoral and tibial resection instruments, which enable surgical resection correctly aligned in the coronal and sagittal planes without damage to the IM canal and associated risks, such as blood loss. Fang et al¹⁰ showed that a well-positioned TKA leads to good clinical results and that malalignment is a predictor for TKA revision surgery (varus knees failed primarily by medial tibia collapse, whereas valgus knees failed from ligament instability).

The purpose of the present study was to compare, in terms of blood loss and implant alignment in the coronal and sagittal planes, a new generation of smart EM cutting guides with the conventional technique based on the use of IM guides. The hypothesis was that the EM system, by avoiding the medullary breaching, would result in less blood loss and less hospitalization stay due to the better postoperative clinical status and hemoglobin values, while ensuring equal or even higher accuracy in the alignment of the femoral implant.

Methods

Study Design

The authors conducted a comparative multicenter randomized study on patients undergoing TKA at two independent centers in which one surgeon for center performed all the surgical procedures. The ethics committees of each center approved the study in advance.

Patients

The inclusion criteria were: patients undergoing primary TKA and age between 60 and 75 years; exclusion criteria were: pathological conditions affecting the ipsilateral knee or ankle, previous surgery on the ipsilateral hip or ankle, intra- or extra-articular deformities limiting the use of the regular IM guide, systemic disease affecting the hemoglobin levels in blood, or the patients' erythropoietic potential, and anticoagulation or antiplatelet therapy.

Thirty-six patients were enrolled (18 patients in each unit). After obtaining a written informed consent, patients were randomly assigned a randomization online software (www.randomization.com) to either of the two study groups 18 patients underwent TKA using the novel smart EM cutting guide (Perseus, Orthokey Italia, Italy; group A) while in the other 18 patients the conventional IM guide was used (group B).

Interventions

The same surgical approach (parapatellar medial capsulotomy) under tourniquet was used in the two groups. Both centers used the same implant (Attune; Depuy, Warsaw, IN, U.S.A.) and all the implants were cemented. Balance of space in flexion and extension (gap balancing technique) was used in both groups.

In the experimental group (A), the femoral distal resection was performed using the EM system based on inertial sensors, while in the control group (B) the femoral distal resection was performed using conventional IM instruments.

The Perseus EM guide consists of a cutting guide assembled with a disposable device containing inertial sensors that communicate via Bluetooth with a tablet placed outside the surgical field (► Fig. 1). After performing the surgical approach, Perseus is fixed to the bone segment close to the resection area (► Fig. 2). To determine the anatomical reference axes, the angular relationship between the electronic pod of the digitizer

Fig. 1 Perseus cutting guide.

Fig. 2 Perseus fixed to bone segment close to resection area.

and the bone reference is registered by the system through movements of the hip: abduction-adduction and flexion-extension. This function activates the inertial mechanism, which then identifies the femoral mechanical axis. Starting from this axis, the triaxial inertial device directs the surgeon on how to align the cutting guide for a correct performance of the resection; plane control was on frontal and sagittal planes (► Fig. 3).

The two groups of patients underwent the same postoperative protocol for rehabilitation, antithromboembolic prophylaxis (enoxaparin 4,000 IU/die) and pain medications.

Outcome Measurements

Preoperative and postoperative (at discharge) nonweight-bearing X-rays were acquired in anteroposterior (AP), lateral,

Fig. 3 Surgical image of femoral cut using extramedullary guide.

Table 1 Baseline characteristics of patients of the two groups

Group	Age (y)	Sex (m/f)	Limb preop alignment (degree)	Hemoglobin (g/dL)
A (EM)	71 ± 5	10/8	4.0 ± 2.5 (varus)	12.0 ± 1.7
B (IM)	72 ± 6	6/12	4.6 ± 4.1 (varus)	12.7 ± 1.4

Abbreviations: EM, extramedullary; IM, intramedullary; preop, preoperative.

and full-limb views, taking care to position the limb in a neutral position. Two independent investigators measured the angles between the prosthetic femoral component and the femoral mechanical axis on the frontal plane and the anterior cortex on the lateral plane. Hemoglobin values were recorded during hospitalization and preoperative values were compared with the minimum postoperative value obtained. Lastly, length of stay (LOS) in hospital was recorded.

Statistical Analysis

Due to the sample size and nonuniform data distribution, the Kruskal-Wallis test was used to compare differences between the two groups in terms of blood loss and implant alignment. Sample size calculation was based on the existing literature, which shows that the average and the standard deviation of the variation of hemoglobin following TKA performed with the conventional IM technique are equal to 3.5 g/dL and 1.2 g/dL, respectively. Assuming that the alternative treatment produces at least a reduction of 1 g/dL hemoglobin level and assuming an α level equal to 0.05 and a power of at least 80%, the minimum number of cases per group required was 18 (total number: 36).

Results

► **Table 1** shows complete baseline data. No differences were observed in terms of age, preoperative limb alignment, or preoperative hemoglobin level in two groups.

Estimate of blood loss showed differences between the two groups (► **Table 2**). The hemoglobin difference was significantly lower in the EM group ($p = 0.0371$).

The results of the alignment of the prosthetic femoral implant are reported in ► **Table 3**. Perpendicularity of the distal resection to the femoral mechanical axis was comparable in the two groups. All patients in the EM cohort had a final alignment within 3 degrees of range on the frontal view,

Table 2 Hemoglobin variation in the two groups

Group	Hemoglobin (g/dL)		
	Preop (min)	Postop (min)	Variation
A (EM)	12.0 ± 1.4	9.3 ± 0.9	2.4 ± 1.3 ^a
B (IM)	12.7 ± 1.7	9.5 ± 1.3	3.3 ± 1.2

Abbreviations: EM, extramedullary; IM, intramedullary; preop, preoperative; postop, postoperative.

^aSignificant difference.

Table 3 Results related to angular deviation of the prosthetic femoral implant

Group	Angular deviation (degree)	
	Frontal plane	Sagittal plane
A (EM)	0.6 ± 1.1 (varus)	-0.6 ± 1.9 (extension)
B (IM)	0.7 ± 2.8 (varus)	2.2 ± 4.4 (extension)

Abbreviations: EM, extramedullary; IM, intramedullary.

while four patients (22%) in the IM cohort had a final alignment exceeding 3 degrees of range (►Fig. 4). Lastly, the mean LOS in hospital showed no significant difference between groups, though group (A) showed 1.6 days less compared with group (B) (7.7 ± 3.1 and 9.3 ± 3.9 days for group A and B, respectively).

Discussion

The main finding of this study was that blood loss was lower in patients who underwent TKA with the EM guide, compared with the conventional IM technique.

Blood loss during TKA is an important issue that is often underestimated. Patients who undergo this type of surgery are usually over the age of 60 years and concomitant patho-

logical conditions such as hypertension, diabetic disorders, and heart diseases are often observed.¹¹ Less blood loss may help to reduce the intraoperative and postoperative risks. Furthermore, significant blood loss may require a blood transfusion, which carries its own risks including infection, abnormal immunological responses, febrile transfusion reactions, and allergic reactions.¹²

This study found that variation in the hemoglobin value during hospitalization was of 2.4 ± 1.3 in group (A) and 3.3 ± 1.2 in group (B). This latter difference was statistically significant. These results are comparable with those of other studies^{4,5} that showed blood loss could be reduced when the medullary canal was not breached during TKA. Jeon et al¹² compared the EM and conventional IM femoral alignment guide systems for bleeding and transfusion rate after TKA and demonstrated that the EM femoral alignment guide technique was advantageous in reducing the volume of blood drained and the allogeneic transfusion rate. In a similar setup, Kandel et al⁵ reported a significant lower average blood drainage volume and transfusion rate using EM guide as well. Using a navigation system to align the femoral component, Kalairajah et al⁴ reported significantly lower hemoglobin loss compared with the conventional procedure.

Implant positioning in the EM group was comparable to IM instrumentation in the final alignment but showed a reduced distribution within the average value, especially in the sagittal plane. On the frontal plane, no patients operated on with EM guide were outside 3 degrees of alignment range, while 22% in the IM cohort had a final alignment outside 3 degrees of range. Although recent studies^{13,14} questioned the role of overall TKA postoperative alignment in implant survival, most of the authors^{1,2,10,15} still advocate placing the implant within 3 degrees of a neutral mechanical axis. Fang et al,¹⁰ in their retrospective study, based on 6,070 TKA patients looking for implant survival in well-aligned knees compared with that of the outliers (more than 3 degrees of valgus or varus alignment) reported that well-aligned TKAs has a longer survival than ones placed in varus or valgus alignment. Wear of the polyethylene tibial bearing surface is a leading cause of failure in knee replacements performed in the last decade due to limb malalignment, patient-related, and implant-related factors.¹⁶

In the present study, EM guide demonstrated proper distal cut alignment with respect to the femoral mechanical axis on the frontal plane. Patients operated with EM guide were all inside the alignment range (100%), whereas the standard alignment in the IM group was observed in 78% of the cases. Nam et al⁷ using a different inertial-based EM guide system reported no significant difference in the number of TKAs within 3 degrees of a neutral mechanical alignment compared with a conventional IM system in a cohort of 94 patients. However, a trend was seen in the EM cohort toward a higher percentage of implants being in the correct alignment range (89.4% vs. 74.5%). Scuderi et al⁶ described the results of another EM inertial-based guide. The authors included 25 patients and reported results from the first 14. The femoral mechanical axis showed an average accuracy of 0.4 ± 0.95 between the confirmed bone surface and the postoperative computed tomography (CT).

Fig. 4 Postoperative X-ray: proper alignment on the frontal and sagittal plane.

A clear advantage of the smart EM cutting guides is the easy to use and the shorter surgical time compared with CAS. Navigation systems is more complex to setup compared with the EM guide as only one pin is required to close the system and the acquisition process is quick and easy. Goh et al¹⁷ included 76 patients and compared the results of the EM inertial-based guide and CAS. They reported no significant difference in clinical outcomes between the two groups. The duration of surgery, however, was significantly longer in the CAS group (101 ± 11 minutes), compared with the EM group (84 ± 21 minutes). The percentage of patients with an alignment within 3 degrees of a neutral mechanical axis for the EM group was 92.1%, which was comparable to the CAS group and to other studies of large-console CAS systems.^{11,18–21} The authors found reproducible accuracy in component alignment with inertial-based femoral resection.

Another crucial aspect in the positioning of the femoral component is flexion in the sagittal plane. In the present study, the EM cutting guide proved better in terms of flexion of the femoral component, but again this difference was not significant. Kim et al¹⁵ found a connection between sagittal femoral component alignment and early component failure. Out of 3,048 knees investigated at 15.8 years, 3.3% of flexed required revision while no neutrally aligned femoral implants required revision. Maderbacher et al² found that IM alignment rods caused a mean sagittal alignment of the cutting block of 4.4 degrees of flexion in relation to the mechanical femoral axis. In only 25% of the observed cases, a flexion within a range of the desired value of 0 to 3 degrees was obtained. The authors concluded that exact sagittal femoral component alignment could only be achieved by navigation devices. In the current study there was no significant difference between EM and the IM cutting guide; however, once again EM guide tended to perform better.

The present study found that, using EM guide, fewer days of hospitalization were necessary compared with patients who underwent TKA using conventional IM guide, albeit this difference was not statistically significant.

The present study has some limitations. First, the choice of using traditional X-rays in place of CT implies possible measurement errors due to limb rotation. Standing AP hip-to-ankle radiographs and lateral knee-to-ankle radiographs were used. It is possible that CT is a more accurate radiologic method, as radiographs can be subject to rotational variations. However, CT has several disadvantages, including radiation exposure and metallic artifacts, and hip-to-ankle radiographs are easily obtained in the routine follow-up, avoiding an increase in the costs of the study. Radiographers were trained to obtain consistent films, and radiographs were repeated if malrotation was detected. Further, it has been shown that radiographs are reproducible for assessing implant placement^{22,23} and have less exposition to radiation. The second limitation was the low number of patients and the lack of follow-up. Lastly, it is difficult to draw conclusions from this study regarding the “learning curve” that is required in the use of the accelerometer-based navigation device. The inclusion period took a considerable amount of time, and surgeons in the study would

often go for several months without using this technique. However, a previous cadaveric study has demonstrated a rapid learning curve with inertial-based cutting guides.¹⁹

In conclusion, the inertial-based EM guide system could be useful to decrease blood loss in TKA and to obtain better postoperative clinical conditions and shorter hospital stay compared with conventional IM guide systems. This device tended to have better results even in terms of accuracy of implant alignment, but these differences were not significant. Future studies are needed to compare the performance of Perseus with other inertial-based systems, which are generally easier to use than navigation systems and less invasive than conventional IM guide systems.

Conflict of Interest
None declared.

References

- Ritter MA, Davis KE, Meding JB, Pierson JL, Berend ME, Malinzak RA. The effect of alignment and BMI on failure of total knee replacement. *J Bone Joint Surg Am* 2011;93(17):1588–1596
- Maderbacher G, Schaumburger J, Baier C, et al. Appropriate sagittal femoral component alignment cannot be ensured by intramedullary alignment rods. *Knee Surg Sports Traumatol Arthrosc* 2016;24(08):2453–2460
- Kim YH. Incidence of fat embolism syndrome after cemented or cementless bilateral simultaneous and unilateral total knee arthroplasty. *J Arthroplasty* 2001;16(06):730–739
- Kalairajah Y, Simpson D, Cossey AJ, Verrall GM, Spriggins AJ. Blood loss after total knee replacement: effects of computer-assisted surgery. *J Bone Joint Surg Br* 2005;87(11):1480–1482
- Kandel L, Vasili C, Kirsh G. Extramedullary femoral alignment instrumentation reduces blood loss after uncemented total knee arthroplasty. *J Knee Surg* 2006;19(04):256–258
- Scuderì GR, Fallaha M, Masse V, Lavigne P, Amiot LP, Berthiaume MJ. Total knee arthroplasty with a novel navigation system within the surgical field. *Orthop Clin North Am* 2014;45(02):167–173
- Nam D, Cody EA, Nguyen JT, Figgie MP, Mayman DJ. Extramedullary guides versus portable, accelerometer-based navigation for tibial alignment in total knee arthroplasty: a randomized, controlled trial: winner of the 2013 HAP PAUL award. *J Arthroplasty* 2014;29(02):288–294
- Nam D, Jerabek SA, Haugom B, Cross MB, Reinhardt KR, Mayman DJ. Radiographic analysis of a hand-held surgical navigation system for tibial resection in total knee arthroplasty. *J Arthroplasty* 2011;26(08):1527–1533
- Nam D, Nawabi DH, Cross MB, Heyse TJ, Mayman DJ. Accelerometer-based computer navigation for performing the distal femoral resection in total knee arthroplasty. *J Arthroplasty* 2012;27(09):1717–1722
- Fang DM, Ritter MA, Davis KE. Coronal alignment in total knee arthroplasty: just how important is it? *J Arthroplasty* 2009;24(6, Suppl):39–43
- Heidari B. Knee osteoarthritis prevalence, risk factors, pathogenesis and features: Part I. *Caspian J Intern Med* 2011;2(02):205–212
- Jeon SH, Kim JH, Lee JM, Seo ES. Efficacy of extramedullary femoral component alignment guide system for blood saving after total knee arthroplasty. *Knee Surg Relat Res* 2012;24(02):99–103
- Parratte S, Pagnano MW, Trousdale RT, Berry DJ. Effect of postoperative mechanical axis alignment on the fifteen-year survival of modern, cemented total knee replacements. *J Bone Joint Surg Am* 2010;92(12):2143–2149
- Sikorski JM. Alignment in total knee replacement. *J Bone Joint Surg Br* 2008;90(09):1121–1127

- 15 Kim YH, Park JW, Kim JS, Park SD. The relationship between the survival of total knee arthroplasty and postoperative coronal, sagittal and rotational alignment of knee prosthesis. *Int Orthop* 2014;38(02):379–385
- 16 Collier MB, Engh CA Jr, McAuley JP, Engh GA. Factors associated with the loss of thickness of polyethylene tibial bearings after knee arthroplasty. *J Bone Joint Surg Am* 2007;89(06):1306–1314
- 17 Goh GS, Liow MH, Lim WS, Tay DK, Yeo SJ, Tan MH. Accelerometer-based navigation is as accurate as optical computer navigation in restoring the joint line and mechanical Axis after total knee arthroplasty: a prospective matched study. *J Arthroplasty* 2016;31(01):92–97
- 18 Mason JB, Fehring TK, Estok R, Banel D, Fahrbach K. Meta-analysis of alignment outcomes in computer-assisted total knee arthroplasty surgery. *J Arthroplasty* 2007;22(08):1097–1106
- 19 Nam D, Dy CJ, Cross MB, Kang MN, Mayman DJ. Cadaveric results of an accelerometer based, extramedullary navigation system for the tibial resection in total knee arthroplasty. *Knee* 2012;19(05):617–621
- 20 Nam D, Weeks KD, Reinhardt KR, Nawabi DH, Cross MB, Mayman DJ. Accelerometer-based, portable navigation vs imageless, large-console computer-assisted navigation in total knee arthroplasty: a comparison of radiographic results. *J Arthroplasty* 2013;28(02):255–261
- 21 Sparmann M, Wolke B, Czupalla H, Banzer D, Zink A. Positioning of total knee arthroplasty with and without navigation support. A prospective, randomised study. *J Bone Joint Surg Br* 2003;85(06):830–835
- 22 Duparc J, Massare C. [Radiological measurement of the angular deviation of the knee in the frontal plane]. *Ann Radiol (Paris)* 1967;10(09):635–656
- 23 Spencer JM, Chauhan SK, Sloan K, Taylor A, Beaver RJ. Computer navigation versus conventional total knee replacement: no difference in functional results at two years. *J Bone Joint Surg Br* 2007;89(04):477–480