Mini Review

Biosensor Designs for Platelet-derived Microparticles Analysis

Jyotsna Kailashiya

Department of Biochemistry, Institute of Medical Sciences,
Banaras Hindu University, Varanasi, UP.

ABSTRACT

Platelet-derived microparticles (PMPs) are often used as marker of platelet activation and their count in blood has been found to be significantly associated with many diseases like myocardial infarction, stroke, venous thrombo-embolism etc. PMPs have been proposed as potential biomarkers for these conditions. Biosensors are newer analytical tools, being developed for convenient and cost effective analysis. For PMPs analysis, biosensors offer many advantages over conventional analysis techniques. This mini review compiles designs and techniques of reported biosensors based on antibody capturing for analysis of PMPs.

Keywords: Platelet-derived microparticles, biosensors, biomarkers, electrochemical analysis, myocardial infarction.

Introduction

Platelets release extracellular vesicles (EVs) in blood. Vesicles originating from cell membrane budding are known as microparticles while smaller vesicles originating from intracellular multivesicular bodies are called exosomes. Platelet-derived microparticles (PMPs) are often used as marker of platelet activation (1, 2). PMPs are most abundant EVs in human blood (3, 4). PMPs are of heterogeneous in size ranging 0.1 to 1 micron (100- 1000 nm), while exosomes size ranges from ~40-100 nm (3). Ninety percent of platelet EV are below 500 nm in size, majority being in range of 100-250 nm (5). PMPs play significant role in cell to cell communication, homeostasis, angiogenesis and other functions. They are rich in phosphatidylserine, tissue factor and many other receptors and have procoagulant surface and ability to interact with leukocytes and endothelial cells (6-8).

PMPs count in blood has been found to be significantly associated with many acute and chronic diseases like myocardial infarction, stroke, venous thrombo-embolism, preeclampsia, fungal (candida albicans) sepsis, systemic lupus erythematosus (SLE), rheumatoid arthritis, etc. (9, 3, 10-13, 4, 14). Increased number of circulating microparticles can trigger hypercoagulable state, which may lead to thrombo-embolic complications (11).

PMPs play role in haemostasis and are considered highly procoagulant (4, 15). Increased blood PMPs concentration in patients of myocardial infarction and acute coronary syndrome, compared to healthy individuals, has already been reported (16, 10, 17, 15). Thus, PMPs level in blood has been proposed as potential biomarker (18, 10, 17). Multiple techniques for PMPs level estimation and characterization have been reported, including flow cytometry, electron microscopy, nanosight

Correspondence : Dr. Jyotsna Kailashiya, Department of Biochemistry, Institute of Medical Sciences, Banaras Hindu University, Varanasi, UP. Email: jyotsna.kailashiya@gmail.com.

tracking analysis, ELISA, etc.

Biosensors are analytical devices that combine biological components with physico-chemical detectors. Biosensors offer many advantages over conventional analysis techniques like low cost, quick results, miniaturization and easy application (19). Immunosensor is a variety of biosensor where principle of antigen-antibody reaction is utilized for specificity of detection and antibodies are fabricated on sensor platform for analyte capture (20). Novel biosensors are being reported regularly and are gaining popularity in medical diagnostic field (20-22). For convenient analysis of PMPs, biosensors are being developed as newer tools.

Designs of biosensors for PMPs analysis

Our lab was first to design specific antibody based PMPs capture nano-bio-sensor platform. We reported a graphene oxide based biosensor for quantitative estimation of PMPs. Glassy carbon disc electrodes were fabricated with graphene oxide at first step, then PMP specific PAC1 (first Procaspase Activating Compound) antibody was immobilized onto it to make the sensor specific for PMPs. Detection was based on electrochemistry by frequency response analysis (FRA), which detects impedance applied by immobilized PMPs on sensor surface. The results were compared and confirmed by flow cytometry analysis of same PMPs samples. Our biosensor successfully and quantitatively estimated PMPs levels in plasma samples of healthy donors as well as acute myocardial infarction patients. The detection limit of our biosensor was 100 PMPs per μL sample, and linear response was seen upto 7000 PMPs per μL for quantitative detection. This biosensor design proved to be most suitable for quick, cost effective, sensitive and user friendly point of care testing and opened new paradigm for development of biosensor based PMPs detection techniques (10).

In 2017, building upon our PMPs

biosensor technique, Singh *et al.* reported another design of biosensor based on nanosilica-PAC1 antibody, P-selectin antibody and conjugated Horse Radish Peroxidase (HRP) on ITO (Indium Tin oxide) electrode (23). Thionine doped silica nanoparticles were synthesized and applied for this biosensor design, where Nanosilica was used as 3-dimentionals platform for biomolecules encapsulation and thionine was used as mediator of electron transfer. PAC1 antibody was used for capturing PMPs on sensor surface, while HRP tagged P-selectin antibody generated detection signal after reacting with H_2O_2 (hydrogen peroxide). Signal generated by H_2O_2 oxidation through HRP tagged P-selectin was used for estimation of PMPs quantity. H_2O_2 oxidation was detected by cyclic voltametry, where higher concentration of PMPs in sample generates higher peaks in cyclic voltamograms, which in turn successfully detected PMPs numbers in samples. Reported detection limit was 490 PMPs per μL sample, and linear response was seen upto 4080 PMPs per μL (23).

A hybrid technique for simultaneous real time quantification and characterization of PMPs by combining surface plasmon resonance (SPR) and atomic force microscopy (AFM) was also developed and reported as an 'on-chip' NanoBioAnalytical platform in 2017 (24). Authors immobilized CD41 antibody on gold sensor chip to capture PMPs, and then used SPR and AFM to characterize captured PMPs. This technique eliminated size limitation, need for labelling PMPs and complex sample preparation. This technique allowed metrological analysis of captured PMPs and revealed that more than 95% of PMPs were smaller than 300 nm. This method reported analysis of PMPs of size range 30 nm to 1 μm and concentration range 109-1010 per μL (24). Results suggested that this NanoBioAnalytical platform, combining SPR and AFM, is a suitable method for a sensitive, reproducible, label-free characterization and quantification of various microparticles over a wide concentration range (24).

Biosensor based techniques offer many advantages by reducing analysis time, eliminating staining steps, ease of application and requirement of very small sample volume for detection. These biosensors are still under development phase, targeted for user friendly wide spread applications.

References

- Hayasaka K, Moriyama T, Chiba H, Matsuno K (2006). Advancement of platelet activation measurement: focusing on platelet-derived microparticle measurement. *Rinsho Byori* **54(3)**: 250-255.
- Heijnen HF, Schiel AE, Fijnheer R, Geuze HJ, Sixma JJ (1999). Activated platelets release two types of membrane vesicles: microvesicles by surface shedding and exosomes derived from exocytosis of multivesicular bodies and alpha-granules. *Blood* **94(11)**: 3791-3799.
- Nieuwland R, Sturk A (2002). Platelet-derived microparticles. *Platelets*: 255-262.
- Italiano JE Jr, Mairuhu AT, Flaumenhaft R (2010). Clinical relevance of microparticles from platelets and megakaryocytes. *Curr Opin Hematol* **17(6)**: 578-584.
- Aatonen MT, Ohman T, Nyman TA, Laitinen S, Gronholm M, Siljander PR (2014). Isolation and characterization of platelet-derived extracellular vesicles. *J Extracell Vesicles* **3(1)**.
- Morel O, Morel N, Freyssinet JM, Toti F (2008). Platelet microparticles and vascular cells interactions: a checkpoint between the haemostatic and thrombotic responses. *Platelets* **19(1)**: 9-23.
- Vajen T, Mause SF, Koenen RR (2015). Microvesicles from platelets: novel drivers of vascular inflammation. *Thromb Haemost* **114(2)**: 228-236.
- Todorova D, Simoncini S, Lacroix R, Sabatier F, Dignat-George F (2017). Extracellular Vesicles in Angiogenesis. *Circ Res* **120(10)**: 1658-1673.
- Woth G, Tokes-Fuzesi M, Magyarlaci T, Kovacs GL, Vermes I, Muhl D (2012). Activated platelet-derived microparticle numbers are elevated in patients with severe fungal (*Candida albicans*) sepsis. *Ann Clin Biochem* **49**: 554-560.
- Kailashiya J, Singh N, Singh SK, Agrawal V, Dash D (2015). Graphene oxide-based biosensor for detection of platelet-derived microparticles: a potential tool for thrombus risk identification. *Biosens Bioelectron* **65**: 274-280.
- Bucciarelli P, Martinelli I, Artoni A, *et al* (2012). Circulating microparticles and risk of venous thromboembolism. *Thromb Res* **129(5)**: 591-597.
- Campello E, Spiezia L, Radu CM, *et al* (2015). Circulating microparticles in umbilical cord blood in normal pregnancy and pregnancy with preeclampsia. *Thromb Res* **136(2)**: 427-431.
- Fortin PR, Cloutier N, Bissonnette V, *et al* (2016). Distinct Subtypes of Microparticle-containing Immune Complexes Are Associated with Disease Activity, Damage, and Carotid Intima-media Thickness in Systemic Lupus Erythematosus. *J Rheumatol* **43(11)**: 2019-2025.
- Marques FK, Campos FM, Filho OA, Carvalho AT, Dusse LM, Gomes KB (2012). Circulating microparticles in severe preeclampsia. *Clinica chimica acta. Intl J Clin Chem* **414**: 253-258.

15. Sun C, Zhao WB, Chen Y, Hu HY (2016). Higher Plasma Concentrations of Platelet Microparticles in Patients With Acute Coronary Syndrome: a Systematic Review and Meta-analysis. *Can J Cardiol* **32(11)**: 1321-1325.
16. Michelsen AE, Brodin E, Brosstad F, Hansen JB (2008). Increased level of platelet microparticles in survivors of myocardial infarction. *Scand J Clin Lab Invest* **68(5)**: 386-392.
17. Kafian S, Mobarrez F, Wallen H, Samad B (2015). Association between platelet reactivity and circulating platelet-derived microvesicles in patients with acute coronary syndrome. *Platelets* **26(5)**: 467-473.
18. Ayers L, Kohler M, Harrison P, *et al* (2011). Measurement of circulating cell-derived microparticles by flow cytometry: sources of variability within the assay. *Thromb Res* **127(4)**: 370-377.
19. Maduraiveeran G, Sasidharan M, Ganesan V (2018). Electrochemical sensor and biosensor platforms based on advanced nanomaterials for biological and biomedical applications. *Biosens Bioelectron* **103**: 113-129.
20. Cho IH, Lee J, Kim J, *et al* (2018). Current Technologies of Electrochemical Immunosensors: Perspective on Signal Amplification. *Sensors* **18(1)**: 207.
21. Ibau C, Md Arshad MK, Gopinath SCB (2017). Current advances and future visions on bioelectronic immunosensing for prostate-specific antigen. *Biosens Bioelectron* **98**: 267-284.
22. Henares TG, Mizutani F, Hisamoto H (2008). Current development in microfluidic immunosensing chip. *Anal Chim Acta* **611(1)**: 17-30.
23. Singh P, Srivastava S, Chakrabarti P, Singh SK (2017). Nanosilica based electrochemical biosensor: A novel approach for the detection of platelet-derived microparticles. *Sensors and Actuators B: Chemical* **240**: 322-329.
24. Obeid S, Ceroi A, Mourey G, Saas P, Elie-Caille C, Boireau W (2017). Development of a NanoBioAnalytical platform for "on-chip" qualification and quantification of platelet-derived microparticles. *Biosens Bioelectron* **93**: 250-259.