Endoscopic ultrasound-guided drainage of a post-hepatectomy abscess using a lumen-apposing self-expandable metal stent with electrocautery-enhanced delivery system

Fig. 1 Cross-sectional computed tomography image of the abscess (arrow) located between the remnant liver and the stomach.

Fig. 2 Endoscopic ultrasound (EUS) view of the distal flange of the lumen-apposing self-expandable metal stent (arrow) that was deployed completely under EUS guidance.

Fig. 3 Endoscopic view of the stent with drainage of pus.

Video 1

Endoscopic ultrasound-guided drainage using a lumen-apposing fully covered self-expandable metal stent for treatment of an intra-abdominal abscess that developed after laparoscopic hepatectomy.

Endoscopic ultrasound (EUS)-guided drainage of pancreatic fluid collections and of the gallbladder and bile duct after failure of standard procedures is becoming an attractive minimally invasive alternative to percutaneous drainage [1–3]. These procedures have been facilitated by the development of a specifically designed lumen-apposing self-expandable metal stent (LA-SEMS) and electrocautery-enhanced delivery system (Hot Axios; Boston Scientific Corp., Marlborough, Massachusetts, USA), which allows the procedure to be performed in one step, mostly under EUS guidance without fluoroscopy [4]. We report on a patient in whom drainage of an abdominal abscess that developed after laparoscopic hepatectomy was performed using the LA-SEMS, thus sparing him from percutaneous drainage and/or repeat surgery.

A 70-year-old man with alcohol-related liver cirrhosis underwent left hepatec-

tomy as treatment for recurrent hepatocellular carcinoma. A month later, he developed persistent abdominal pain with fever. An abdominal computed tomography scan revealed an abscess located between the remnant right liver lobe and the stomach (● **Fig. 1**). EUS-guided drainage was offered to the patient and he signed an informed consent form. Using a therapeutic linear echoendoscope, the 7-cm collection was located and the presence of interposing vessels was excluded using Doppler. Direct transgastric penetration into the collection using the Hot Axios device was accomplished by applying pure cut cautery, and was followed by EUS-guided deployment of a 10×10-mm LA-SEMS (● **Fig. 2**, ● **Video 1**), with drainage of purulent material (● **Fig. 3**, ● **Video 1**). No complications occurred. The patient was discharged completely asymptotically the following day.

Removal of the stent was performed 4 weeks later, and complete resolution of the abscess was observed.

Intra-abdominal abscess formation after laparoscopic hepatectomy has been reported to occur in up to 10% of cases [5]. The case presented shows that this novel LA-SEMS is safe, easy to use, and highly effective, and should be considered when intra-abdominal abscesses accessible to EUS are detected.

Endoscopy_UCTN_Code_TTT_1AS_2AC

Competing interests: Dr. Larghi is a consultant for Boston Scientific Corp.

**Fabia Attili^{1,2}, Shyam Dang^{1,3},
Mihai Rimbaş^{1,4}, Luca Di Maurizio^{1,2},
Giuseppe Maria Ettorre^{2,5}, Adolfo
Francesco Attili^{2,6}, Alberto Larghi^{1,2}**

- ¹ Digestive Endoscopy Unit, Catholic University, Rome, Italy
- ² Salvator Mundi, International Hospital, Rome, Italy
- ³ University of Arkansas Medical Sciences, Little Rock, Arkansas, United States
- ⁴ Department of Gastroenterology, Colentina Clinical Hospital, Carol Davila University of Medicine, Bucharest, Romania
- ⁵ Division of General Surgery and Liver Transplantation, S. Camillo Hospital, Rome, Italy
- ⁶ Department of Clinical Medicine, Division of Gastroenterology, "Sapienza" University of Rome, Rome, Italy

References

- 1 Rinninella E, Kunda R, Dollhopf M et al. EUS-guided drainage of pancreatic fluid collections using a novel lumen-apposing metal stent on an electrocautery-enhanced delivery system: a large retrospective study (with video). *Gastrointest Endosc* 2015; 82: 1039–1046
- 2 Walter D, Teoh AY, Itoi T et al. EUS-guided gall bladder drainage with a lumen-apposing metal stent: a prospective long-term evaluation. *Gut* 2016; 65: 6–8
- 3 Kunda R, Pérez-Miranda M, Will U et al. EUS-guided choledochoduodenostomy for malignant distal biliary obstruction using a lumen-apposing fully covered metal stent after failed ERCP. *Surg Endosc* 2016. doi: 10.1007/s00464-016-4845-6
- 4 Attili F, Rimbaş M, Galasso D et al. Fluoroless endoscopic ultrasound-guided biliary drainage after failed ERCP with a novel lumen-apposing metal stent mounted on a cautery-tipped delivery system. *Endoscopy* 2015; 47 (Suppl. 01): E619–620
- 5 Matsumata T, Yanaga K, Shimada M et al. Occurrence of intraperitoneal septic complications after hepatic resections between 1985 and 1990. *Surg Today* 1995; 25: 49–54

Bibliography

DOI <http://dx.doi.org/10.1055/s-0042-109051>
Endoscopy 2016; 48: E222–E223
 © Georg Thieme Verlag KG
 Stuttgart · New York
 ISSN 0013-726X

Corresponding author

Alberto Larghi, MD, PhD
 Digestive Endoscopy Unit
 Catholic University
 Largo A. Gemelli 8
 00168, Rome
 Italy
 Fax: +39-06-30156581
alberto.larghi@yahoo.it