




## Development of fatal systemic gas embolism during direct peroral cholangioscopy under carbon dioxide insufflation



**Fig. 1** Images of a hepatolith (white arrow), 20 mm in diameter, in the left intrahepatic bile duct on: **a** computed tomography (CT) scan; **b** magnetic resonance cholangiopancreatography (MRCP).



**Fig. 2** Radiographic images showing: **a** an attempt to extract the stones through the double-balloon enteroscope (DBE) from the distal side of the anastomosis; **b** an ultraslim endoscope, which had replaced the DBE, advanced into the bile duct.




**Fig. 3** Views via the ultraslim endoscope showing: **a** the hepatolith along with pus and mucus in the left intrahepatic bile duct; **b** the proximal bile duct beyond the hepatolith seen at the moment of lithotripsy using a Holmium:YAG laser after mucus and pus had been removed.

Direct peroral cholangioscopy (DPOCS) is a useful and effective technique for diagnosis and therapy of biliary tract disease [1–3]. No instances of a fatal systemic gas embolism developing during DPOCS under carbon dioxide (CO<sub>2</sub>) insufflation have yet been reported.

A 68-year-old woman was admitted to our hospital with a complaint of hepatolithiasis (● Fig. 1). She had undergone Roux-en-Y hepaticojejunostomy for choledochal cysts 34 years previously. We performed DPOCS using a short-type double-balloon enteroscope (DBE), an ultraslim endoscope, and an endoscopic CO<sub>2</sub> regulator (EC-450BI5, EG-580NW, GW-1, respectively; Fujifilm Corp., Tokyo, Japan) while the patient was kept adequately sedated with midazolam.

We planned to perform lithotripsy of the hepatolith using a Holmium:YAG laser. After we had reached the anastomosis using the DBE, attempts to extract the stones through the DBE using balloon or basket catheters failed (● Fig. 2a). We therefore decided to perform DPOCS with an ultraslim endoscope passed through an overtube using a previously described method (● Fig. 2b) [3]. The balloon attached to the overtube remained inflated from the time that we reached the anastomosis until the end of the procedure.

We first confirmed the hepatolith was present (● Fig. 3a). We then prepared the Holmium:YAG laser for lithotripsy for 5 minutes, while we aspirated pus and mucus discharged from the peripheral bile duct near the hepatolith. As we fractured the hepatolith with the Holmium:YAG laser (● Fig. 3b), the patient suddenly went into shock and had a cardiac arrest. Despite immediate cardiomegaly resuscitation and injection of flumazenil, she died. A computed tomography (CT) scan performed during resuscitation revealed multiple gas emboli in the systemic arteries and veins (● Fig. 4). Pathological examination later revealed hepatic abscesses, inflammation surrounding the hepatolith, intravascular gas, and systemic gas emboli [4]. There was no evidence of a patent foramen ovale [5]. The cause of death was systemic gas embolism. We believe aspiration of pus and mucus prior to lithotripsy may have opened a pre-existing biliovenous shunt. Endoscopists should take the possibility of fatal gas embolism into consideration during DPOCS even under CO<sub>2</sub> insufflation. The extent of insufflation should be the absolute minimum required.



**Fig. 4** Computed tomography (CT) scan during resuscitation showing multiple gas emboli in the systemic arteries and veins of: **a** the brain; **b** the heart.

- 4 Donepudi S, Chavalitdhamrong D, Pu L et al. Air embolism complicating gastrointestinal endoscopy: A systematic review. *World J Gastrointest Endosc* 2013; 5: 359–365
- 5 Finsterer J, Stollberger C, Bastovansky A. Cardiac and cerebral air embolism from endoscopic retrograde cholangio-pancreatography. *Eur J Gastroenterol Hepatol* 2010; 22: 1157–1162

#### Bibliography

**DOI** <http://dx.doi.org/10.1055/s-0042-109056>  
*Endoscopy* 2016; 48: E215–E216  
 © Georg Thieme Verlag KG  
 Stuttgart · New York  
 ISSN 0013-726X

Endoscopy\_UCTN\_Code\_CPL\_1AK\_2AI

**Competing interests:** None

**Hiromu Kondo, Itaru Naitoh, Takahiro Nakazawa, Kazuki Hayashi, Yuji Nishi, Shuichiro Umemura, Takashi Joh**

Department of Gastroenterology and Metabolism, Nagoya City University Graduate School of Medical Sciences, Nagoya, Japan

#### References

- 1 Tringali A, Lemmers A, Meves V et al. Intra-ductal biliopancreatic imaging: European Society of Gastrointestinal Endoscopy (ESGE) technology review. *Endoscopy* 2015; 47: 739–753
- 2 Moon JH, Ko BM, Choi HJ et al. Direct peroral cholangioscopy using an ultra-slim upper endoscope for the treatment of retained bile duct stones. *Am J Gastroenterol* 2009; 104: 2729–2733
- 3 Itoi T, Sofuni A, Itokawa F et al. Diagnostic and therapeutic peroral direct cholangioscopy in patients with altered GI anatomy (with videos). *Gastrointest Endosc* 2012; 75: 441–449

#### Corresponding author

**Itaru Naitoh, MD**

Department of Gastroenterology and Metabolism  
 Nagoya City University Graduate  
 School of Medical Sciences  
 1 Kawasumi, Mizuho-cho, Mizuho-ku  
 Nagoya 467-8601  
 Japan  
 Fax: +81-52-8520952  
 inaito@med.nagoya-cu.ac.jp