Development of fatal systemic gas embolism during direct peroral cholangioscopy under carbon dioxide insufflation

Fig. 1 Images of a hepatolith (white arrow), 20 mm in diameter, in the left intrahepatic bile duct on: **a** computed tomography (CT) scan; **b** magnetic resonance cholangiopancreatography (MRCP).

Fig. 2 Radiographic images showing: **a** an attempt to extract the stones through the double-balloon enteroscope (DBE) from the distal side of the anastomosis; **b** an ultraslim endoscope, which had replaced the DBE, advanced into the bile duct.

Fig. 3 Views via the ultraslim endoscope showing: **a** the hepatolith along with pus and mucus in the left intrahepatic bile duct; **b** the proximal bile duct beyond the hepatolith seen at the moment of lithotripsy using a Holmium:YAG laser after mucus and pus had been removed.

Direct peroral cholangioscopy (DPOCS) is a useful and effective technique for diagnosis and therapy of biliary tract disease [1–3]. No instances of a fatal systemic gas embolism developing during DPOCS under carbon dioxide (CO₂) insufflation have yet been reported.

A 68-year-old woman was admitted to our hospital with a complaint of hepatolithiasis (● Fig. 1). She had undergone Roux-en-Y hepaticojejunostomy for choledochal cysts 34 years previously. We performed DPOCS using a short-type double-balloon enteroscope (DBE), an ultraslim endoscope, and an endoscopic CO₂ regulator (EC-450BI5, EG-580NW, GW-1, respectively; Fujifilm Corp., Tokyo, Japan) while the patient was kept adequately sedated with midazolam.

We planned to perform lithotripsy of the hepatolith using a Holmium:YAG laser. After we had reached the anastomosis using the DBE, attempts to extract the stones through the DBE using balloon or basket catheters failed (● Fig. 2a). We therefore decided to perform DPOCS with an ultraslim endoscope passed through an overtube using a previously described method (● Fig. 2b) [3]. The balloon attached to the overtube remained inflated from the time that we reached the anastomosis until the end of the procedure.

We first confirmed the hepatolith was present (● Fig. 3a). We then prepared the Holmium:YAG laser for lithotripsy for 5 minutes, while we aspirated pus and mucus discharged from the peripheral bile duct near the hepatolith. As we fractured the hepatolith with the Holmium:YAG laser (● Fig. 3b), the patient suddenly went into shock and had a cardiac arrest. Despite immediate cardiomegaly resuscitation and injection of flumazenil, she died. A computed tomography (CT) scan performed during resuscitation revealed multiple gas emboli in the systemic arteries and veins (● Fig. 4). Pathological examination later revealed hepatic abscesses, inflammation surrounding the hepatolith, intravascular gas, and systemic gas emboli [4]. There was no evidence of a patent foramen ovale [5]. The cause of death was systemic gas embolism. We believe aspiration of pus and mucus prior to lithotripsy may have opened a pre-existing biliovenous shunt. Endoscopists should take the possibility of fatal gas embolism into consideration during DPOCS even under CO₂ insufflation. The extent of insufflation should be the absolute minimum required.

Fig. 4 Computed tomography (CT) scan during resuscitation showing multiple gas emboli in the systemic arteries and veins of: **a** the brain; **b** the heart.

Endoscopy_UCTN_Code_CPL_1AK_2AI

Competing interests: None

Hiromu Kondo, Itaru Naitoh, Takahiro Nakazawa, Kazuki Hayashi, Yuji Nishi, Shuichiro Umemura, Takashi Joh

Department of Gastroenterology and Metabolism, Nagoya City University Graduate School of Medical Sciences, Nagoya, Japan

References

- 1 *Tringali A, Lemmers A, Meves V et al.* Intra-ductal biliopancreatic imaging: European Society of Gastrointestinal Endoscopy (ESGE) technology review. *Endoscopy* 2015; 47: 739–753
- 2 *Moon JH, Ko BM, Choi HJ et al.* Direct peroral cholangioscopy using an ultra-slim upper endoscope for the treatment of retained bile duct stones. *Am J Gastroenterol* 2009; 104: 2729–2733
- 3 *Itoi T, Sofuni A, Itokawa F et al.* Diagnostic and therapeutic peroral direct cholangioscopy in patients with altered GI anatomy (with videos). *Gastrointest Endosc* 2012; 75: 441–449
- 4 *Donepudi S, Chavalitdhamrong D, Pu L et al.* Air embolism complicating gastrointestinal endoscopy: A systematic review. *World J Gastrointest Endosc* 2013; 5: 359–365
- 5 *Finsterer J, Stollberger C, Bastovansky A.* Cardiac and cerebral air embolism from endoscopic retrograde cholangio-pancreatography. *Eur J Gastroenterol Hepatol* 2010; 22: 1157–1162

Bibliography

DOI <http://dx.doi.org/10.1055/s-0042-109056>
Endoscopy 2016; 48: E215–E216
 © Georg Thieme Verlag KG
 Stuttgart · New York
 ISSN 0013-726X

Corresponding author

Itaru Naitoh, MD

Department of Gastroenterology and Metabolism
 Nagoya City University Graduate
 School of Medical Sciences
 1 Kawasumi, Mizuho-cho, Mizuho-ku
 Nagoya 467-8601
 Japan
 Fax: +81-52-8520952
inaito@med.nagoya-cu.ac.jp