Endoscopic ultrasonography-guided antegrade stenting combined with hepaticogastrostomy/hepaticojunostomy using ultraslim instruments

► **Fig. 1** The VisiGlide 2 (0.025 inch, angled type; Olympus, Tokyo, Japan) has enhanced tip flexibility and provides the same stiffness as a conventional 0.035-inch guidewire.

► **Fig. 2** A tapered endoscopic retrograde cholangiopancreatography catheter (01 20 21 1; 4.8-Fr tip diameter, 6.9-Fr shaft diameter; MTW Endoskopie, Düsseldorf, Germany) passed over a 0.025-inch guidewire (upper image) and a novel ultraslim uncovered self-expandable metal stent (BileRush Selective; 5.7 Fr, 10-mm diameter, 185-cm long; Piolax Medical Devices, Kanagawa, Japan) (lower image).

Techniques for endoscopic ultrasonography (EUS)-guided biliary drainage (EUS-BD) have been developed, and EUS-guided antegrade stenting (EUS-AGS) and EUS-guided hepaticogastrostomy (EUS-

HGS)/hepaticojunostomy (HJS) are suitable for gastric outlet obstruction (GOO) or surgically altered anatomy. EUS-AGS alone carries the potential risk of causing bile leakage from a fistula; however, EUS-AGS in combination with EUS-HGS or EUS-HJS appears safer, as it can reduce the risk of a bile leak [1, 2].

We present two patients who underwent EUS-HGS or EUS-HJS combined with EUS-AGS using ultraslim instruments. Patient #1 was a 62-year-old woman who had undergone a previous total gastrectomy for gastric cancer and later developed obstructive jaundice. First, a B3 branch was punctured using a 19G needle via a transjejunal approach, and a 0.025-inch guidewire (VisiGlide 2; Olympus, Tokyo, Japan) (► **Fig. 1**) was placed. Next, a tapered endoscopic retrograde cholangiopancreatography (ERCP) catheter (01 20 21 1; MTW Endoskopie, Düsseldorf, Germany) (► **Fig. 2**) was used to dilate the fistula, following successful passage of the guidewire through the stricture. EUS-AGS was then performed using a novel ultraslim uncovered self-expandable metal stent (SEMS; BileRush Selective; 5.7 Fr, 10-mm diameter; Piolax Medical Devices, Kanagawa, Japan)

(► **Fig. 2**). Finally, a novel 7-Fr plastic stent (TYPE-IT stent; Gadelius Medical Co. Ltd., Tokyo, Japan) [3] (► **Fig. 3**) was placed to create an EUS-HJS (► **Fig. 4**; ► **Video 1**).

Patient #2 was a 68-year-old man with GOO caused by gastric cancer who developed obstructive jaundice. EUS-AGS and EUS-HGS were performed as described

► VIDEO 1

► Video 1: (Patient #1) Endoscopic ultrasonography (EUS)-guided antegrade stenting using a novel 5.7-Fr ultraslim uncovered metal stent, and EUS-guided hepaticojunostomy using a novel 7-Fr dedicated plastic stent.

► VIDEO 2

► Video 2: (Patient #2) Endoscopic ultrasonography (EUS)-guided antegrade stenting using a novel 5.7-Fr ultraslim uncovered metal stent, and EUS-guided hepaticogastrostomy using a novel 7-Fr dedicated plastic stent.

► **Fig. 3** A novel plastic stent (TYPE-IT stent; Gadelius Medical Co. Ltd., Tokyo, Japan) can be retracted with a lasso, and has a 7-Fr diameter (5-Fr inner catheter-tip diameter), with total length of 20 cm, effective length of 15 cm, single-pigtail type (proximal end), and four flanges (two at the distal end and two at the proximal end).

► **Fig. 4** (Patient #1) Radiograph showing an endoscopic ultrasonography-guided hepaticojunostomy with a dedicated 7-Fr plastic stent and antegrade stenting with a 5.7-Fr uncovered metal stent.

► **Fig. 5** (Patient #2) Radiograph showing an endoscopic ultrasonography-guided hepaticogastrostomy with a dedicated 7-Fr plastic stent and antegrade stenting with a 5.7-Fr uncovered metal stent.

above (► **Fig. 5**; ► **Video 2**). There were no complications in either case. A covered SEMS (CSEMS) is commonly used to prevent bile leaks in EUS-HGS/HJS. A long partially covered SEMS (PCSEMS; ≥ 10 mm) can be used to prevent stent migration [4]. However, the thicker delivery system (8.5 Fr) with this long PCSEMS and the cost of two metal stents are of concern. In particular, minimum fistula dilation should be performed during EUS-BD. Therefore, EUS-AGS and EUS-HGS/HJS using various ultraslim instruments (7 Fr or less) in combination

can facilitate the procedure and minimize the potential for bile leakage.

Endoscopy_UCTN_Code_TTT_1AS_2AD

Competing interests

A novel ultraslim uncovered metal stent (BileRush Selective; 5.7 Fr, 8-mm/10-mm diameter, 185-cm long) has been developed through collaborative research between Dr. Kawakami and Piolax Medical Devices, Kanaga-

wa, Japan. Dr. Kawakami is a consultant and gives lectures for the Piolax Medical Devices and for Olympus, Tokyo, Japan. Dr. Kubota has no competing interests to declare.

The Authors

Hiroshi Kawakami, Yoshimasa Kubota

Department of Gastroenterology and Hepatology, Faculty of Medicine, University of Miyazaki and Center for Digestive Disease, University of Miyazaki Hospital, Miyazaki, Japan

Corresponding author

Hiroshi Kawakami, MD, PhD

Department of Gastroenterology and Hepatology, Faculty of Medicine, University of Miyazaki, Center for Digestive Disease, University of Miyazaki Hospital, 5200, Kihara, Kiyotake, Miyazaki 889-1692, Japan
Fax: +81-985-859802
hiropon@med.miyazaki-u.ac.jp

References

- [1] Ogura T, Masuda D, Imoto A et al. EUS-guided hepaticogastrostomy combined with fine-gauge antegrade stenting: a pilot study. *Endoscopy* 2014; 46: 416–421
- [2] Kumbhari V, Tieu AH, Khashab MA. EUS-guided biliary drainage made safer by a combination of hepaticogastrostomy and antegrade transpapillary stenting. *Gastrointest Endosc* 2014; 81: 1015–1016
- [3] Umeda J, Itoi T, Tsuchiya T et al. A newly designed plastic stent for EUS-guided hepaticogastrostomy: a prospective preliminary feasibility study (with videos). *Gastrointest Endosc* 2015; 82: 390–396.e2
- [4] Nakai Y, Isayama H, Yamamoto N et al. Safety and effectiveness of a long, partially covered metal stent for endoscopic ultrasound-guided hepaticogastrostomy in patients with malignant biliary obstruction. *Endoscopy* 2016; 48: 1125–1128

Bibliography

DOI <http://dx.doi.org/10.1055/s-0043-101225>
Endoscopy 2017; 49: E88–E89
© Georg Thieme Verlag KG
Stuttgart · New York
ISSN 0013-726X