Figure 1 The irrigation wiper-knife. **a** The irrigation tube (4) and needle-knife (1) were glued to the exterior surface over the hole, but only the needle-knife was intersected by and fixed with the handling wire (2). **b** The schematic representation shows more clearly how the needle-knife moves like a windshield wiper (double-headed red arrow).

Endoscopic submucosal dissection techniques require special skills and the procedure can be very prolonged [1,2]. In response these drawbacks, we have designed a new dissection knife, the “wiper-knife”, which simplifies the procedure.

We recently developed a hood-knife that facilitates endoscopic submucosal dissection by sliding the hood-knife with coagulation current along the muscle layer [2]. However, we considered that it was essential to be able to perform endoscopic

Figure 2 Schematic representation of the endoscopic submucosal dissection procedure using the wiper-knife.

submucosal dissection more mechanically. The wiper-knife was made by installing a needle-knife in exchange for the snare forceps we used in the hood-knife. The needle-knife was crossed by and fixed with a handling wire, which was inserted through tubes at either side of the hood.

The endoscopic submucosal dissection procedure using the wiper-knife is carried out as follows (Figure 2): a grasping forceps is passed through an accessory channel and used to push the lesion away from the muscle layer; submucosal exfoliation is then performed by moving the wiper-knife like a windshield wiper with coagulation current on the muscle layer to separate the submucosa from the muscle layer. We resected three specimens in animal models (pigs) using this wiper-knife. The average diameter of the resected specimens was 30 mm.

Endoscopic submucosal dissection can be safely and easily carried out under direct vision using this method. Another advantage of the wiper-knife is that endoscopic submucosal dissection and endoscopic hemostasis can be carried out while simultaneously applying adequate irrigation [2–6].

Endoscopy_UCTN_Code_TTT_1AO_2AG

K. Kume, M. Yamasaki, K. Kanda, I. Yoshikawa, M. Otsuki

Third Department of Internal Medicine, School of Medicine, University of Occupational and Environmental Health, Kitakyusyu, Japan.

References

- 1 Muto M, Miyamoto S, Hosokawa A et al. Endoscopic mucosal resection in the stomach using the insulated-tip needle-knife. *Endoscopy* 2005; 37: 178–182
- 2 Kume K, Yamasaki M, Kanda K et al. Endoscopic submucosal dissection using a novel irrigation hood-knife. *Endoscopy* 2005; 37: 1030–1031
- 3 Kume K, Yamasaki M, Yamasaki T et al. Endoscopic hemostatic treatment under irrigation for upper GI hemorrhage: a comparison of one third and total circumference transparent end hoods. *Gastrointest Endosc* 2004; 59: 712–716
- 4 Kume K, Yoshikawa I, Otsuki M. Endoscopic treatment of upper GI hemorrhage with a novel irrigating hood attached to the endoscope. *Gastrointest Endosc* 2003; 57: 732–735
- 5 Kume K, Yamasaki M, Kubo K et al. EMR of upper GI lesions when using a novel soft, irrigation, prelooped hood. *Gastrointest Endosc* 2004; 60: 124–128
- 6 Kume K, Yamasaki M, Kanda K et al. Endoscopic procedure under irrigation. *Dig Endosc* 2005; 17: 241–245

Corresponding author

M. Otsuki, M.D.

Third Department of Internal Medicine, School of Medicine
University of Occupational and Environmental Health, 1-1 Iseigaoka, Yahatanishi-ku, Kitakyusyu 807-8555 Japan

Fax: +81-93-692-0107

Email: mac-otsk@med.uoeh-u.ac.jp