Endoscopic Mucosectomy of the Esophagus with a "Shark-Tooth" Polypectomy Snare

Simplified endoscopic resection of the esophageal mucosa using a cap-fitted panendoscope had been described by Inoue et al. (1). We report here on the use of a new type of polypectomy snare in the course of mucosectomy.

A 43-year-old woman presented with weight loss. Panendoscopy demonstrated a reddish flat area with a diameter of 7 mm in the esophagus, 28 cm from the incisors. The area was stained endoscopically with a 1% toluidine blue solution. Both the visible lesion and three other areas near it, each 5–6 mm in diameter and lying far apart, were stained. The largest one was removed completely by biopsy; the histology confirmed squamous-cell cancer. Endoscopic ultrasonography of the esophagus was negative. Since we considered that the lesions in the esophagus represented a multifocal intraepithelial cancer, we therefore decided to carry out an endoscopic mucosectomy in order to remove the lesions. At the endoscopic examination, the mucosal area was slightly hyperemic where the biopsy had been taken, but it did not stain with toluidine blue solution. The other three lesions became visible once again only after staining. Using an endoscopic injector, a physiological saline solution containing epinephrine 1:10,000 was injected underneath the lesions. The protruding polyp-like lesions were then grasped one by one with the "shark-tooth" Olympus SB16-LE polypectomy snare (Figure 1a), and removed by high-frequency diathermic cutting (Figure 1b). The histological examination of the lesions showed moderate dysplasia, and in one lesion also hyperkeratosis. A control panendoscopy and esophageal biopsy one month later were negative.

In our opinion, the use of toluidine blue solution prior to mucosectomy is advantageous, because the long-lasting staining of the lesions makes mucosectomy easier to perform. The use of the "shark-tooth" polypectomy snare made it easier to grasp the mucosal lesions and pull them into the esophageal lumen (Figure 2). The use of this snare is recommended when a transparent endoscope cap is not available, or if the diameter of the lesion is too large.

I. Szántó¹, J. Kiss¹, A. Vörös¹, Á. Altörjay¹, P. Nagy²

¹ Dept. of Surgery

² Institute of Pathology, Postgraduate Medical School, Budapest, Hungary

Reference

1. Inoue H, Takeshita K, Hori H, et al. Endoscopic mucosal resection with a cap-fitted panendoscope for esophagus, stomach, and colon mucosal lesion. *Gastrointest Endosc* 1993; 39: 58–62.

Corresponding Author

I. Szántó, M.D.
Dept. of Surgery
Postgraduate Medical School
Szaboles u 35, P.O. Box 112
1389 Budapest
Hungary

Figure 1: Mucosectomy technique.
a Catching the lesion with the polypectomy snare.

Figure 1b Site of the mucosectomy.

Figure 2: The "shark-tooth" polypectomy snare.