



Figure 1: The working arrangement. The aspiration cannula is connected to a tube which drains fat into the mucous sucker. The suction is maintained by the tubing connected to the other opening in the mucous sucker



Figure 2: The infant mucus sucker showing aspirated fat

An alternative receptacle for fat harvest: The infant mucous sucker

Sir,

Fat grafting is a commonly performed aesthetic procedure. This involves harvesting fat under low pressure using syringe or high pressure using the suction machine.^[1] For harvesting a small amount of fat, a low-pressure system using syringe is good enough. The usual method of harvesting fat graft with 10 or 20 cc syringes becomes tedious and repetitive if a larger amount of fat is

required. The powered suction apparatus is an option, but the aspirated fat goes directly to the suction bottle. One needs to buy a sterile, autoclavable container which is interposed in between the cannula and the machine so that the aspirated fat does not go directly to the suction machine.^[2]

Our modification includes attaching the harvesting suction cannula to a sterile infant mucous sucker via a 2 cc syringe. The other end of this syringe is attached to the suction tubing. After suctioning, the fat gets accumulated in the mucous sucker [Figure 1]. We can then transfer it to the syringes for insertion after the receptacle is full [Figure 2].

The advantage of this modification is that it is easy to set up. It provides a disposable sterile receptacle, which is cheap at Rs. 40/unit. The capacity of the mucous sucker is 20 cc. We can harvest 150 to 200 cc of fat with less

hassle. It is a cheap alternative to the sterile autoclavable and reusable container and is universally available.

Financial support and sponsorship

Nil.

Conflicts of interest

There are no conflicts of interest.

**Satyaswarup Tripathy, Saurabh Rawat,
Jerry R. John, Ramesh Kumar Sharma**

Department of Plastic Surgery, PGIMER, Chandigarh, India

Address for correspondence:

Dr. Satyaswarup Tripathy, Department of Plastic Surgery,
PGIMER, Chandigarh, India.
E-mail: dr.sstripathy@gmail.com

REFERENCES

1. Gir P, Brown SA, Oni G, Kashefi N, Mojallal A, Rohrich RJ. Fat grafting: Evidence-based review on autologous fat harvesting, processing, reinjection, and storage. *Plast Reconstr Surg* 2012;130:249-58.
2. Keck M, Kober J, Riedl O, Kitzinger HB, Wolf S, Stulnig TM, *et al.* Power assisted liposuction to obtain adipose-derived stem cells: Impact on viability and differentiation to adipocytes in comparison to manual aspiration. *J Plast Reconstr Aesthet Surg* 2014;67:e1-8.

This is an open access article distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike 3.0 License, which allows others to remix, tweak, and build upon the work non-commercially, as long as the author is credited and the new creations are licensed under the identical terms.

Access this article online

Quick Response Code:



Website:

www.ijps.org

DOI:

10.4103/0970-0358.182227

How to cite this article: Tripathy S, Rawat S, John JR, Sharma RK. An alternative receptacle for fat harvest: The infant mucous sucker. *Indian J Plast Surg* 2016;49:126-7.