

Letter to Editor

Accuracy of dose calculation algorithms in Eclipse treatment planning system: An update

Dear Editor,

Recently, I read an excellent article entitled “Dose prediction accuracy of anisotropic analytical algorithm and pencil beam convolution algorithm beyond high density heterogeneity interface”^[1] published in South Asian Journal of Cancer. In that study, Rana *et al.*,^[1] concluded that the anisotropic analytical algorithm (AAA) is more accurate than the pencil beam convolution (PBC) algorithm for dose calculations in external beam radiation therapy. This information can serve as one of the factors that can help in selecting AAA vs. PBC for dose calculations in the Eclipse treatment planning system (TPS). While results from Rana *et al.*, are accurate comparing AAA with PBC, I would like to point out that Eclipse TPS has recently implemented new generation dose calculation algorithm called Acuros XB. Since Acuros XB, a Monte Carlo based algorithm, is now widely available for dose calculations, several researchers^[2-5] have investigated the dose prediction accuracy of Acuros XB, and compared the results of Acuros XB with that of AAA measurements, and Monte Carlo simulations. The studies involving Acuros XB showed that Acuros XB is superior to AAA when dose calculations are performed in the presence of heterogeneities. For instance, Bush *et al.*,^[3] showed that, in comparison to the Monte Carlo results, Acuros XB had up to 4.5% difference, whereas AAA results had differences up to 10.2% in lung and up to 17.5% in low-density lung. Similarly, Rana *et al.*,^[4] showed the difference between Acuros XB and measurement results was up to 3.8%, whereas AAA had a larger deviation from the measurement, with difference up to 10.9% beyond the air gap.

Although AAA has been a preferred option over PBC in Eclipse users for the past few years,^[1,6] it must be

noted that Acuros XB is more accurate to use for dose calculations in the Eclipse TPS.^[2-5] Furthermore, the accuracy of dose calculation in radiation therapy is particularly important to avoid the dose underestimation or overestimation. It is recommended that Eclipse users explore an option to calculate the cancer treatment plans using Acuros XB instead of AAA or PBC if Acuros XB is available in their TPS.

Biniam Tesfamicael

Department of Physics, Hampton University, Virginia, USA

Correspondence to: Mr. Biniam Tesfamicael,

E-mail: biniam.yt@gmail.com

References

1. Rana SB. Dose prediction accuracy of anisotropic analytical algorithm and pencil beam convolution algorithm beyond high density heterogeneity interface. South Asian J Cancer 2013;2:26-30.
2. Vassiliev ON, Wareing TA, McGhee J, Failla G, Salehpour MR, Mourtada F. Validation of a new grid-based Blotzmann equation solver for dose calculation in radiotherapy with photon beams. Phys Med Biol 2010;55:581-98.
3. Bush K, Gagne IM, Zavgorodni S, Ansbacher W, Beckham W. Dosimetric validation of Acuros XB with Monte Carlo methods for photon dose calculations. Med Phys 2011;38:2208-21.
4. Rana S, Rogers K. Dosimetric evaluation of Acuros XB dose calculation algorithm with measurements in predicting doses beyond different air gap thickness for smaller and larger field sizes. J Med Phys 2013;38:9-14.
5. Rana S, Rogers K, Lee T, Reed D, Biggs C. Verification and Dosimetric Impact of Acuros XB algorithm for stereotactic body radiation therapy (SBRT) and rapidArc planning for non-small-cell lung cancer (NSCLC) patients. Int J Med Phys Clin Eng Rad Onc 2013;2:6-14.
6. Robinson D. Inhomogeneity correction and the analytic anisotropic algorithm. J Appl Clin Med Phys 2008;9:2786.

Access this article online

Quick Response Code:



Website:

www.sajc.org

DOI:

10.4103/2278-330X.119892