



Coil valve syndrome: a rare complication of percutaneous transhepatic obliteration successfully treated by argon plasma coagulation and double-balloon endoscopy



► **Fig. 1** Cerecyte coil.

Percutaneous transhepatic obliteration (PTO) is now widely used for prophylactic treatment of gastric varices [1]. A straying coil tip in the stomach is sometimes reported as an adverse event [2,3], but there have been no reports of one reaching the small intestine from the stomach. In this case, the migrated coil tip with food residue was shaped like a ball and passed into the jejunum, causing a phenomenon resembling ball valve syndrome [4].

A 70-year-old woman was admitted with epigastric pain. She had been treated for gastric varices by PTO with coils (► **Fig. 1**) 4 years earlier. One of these had migrated into the stomach asymptotically 1 year after PTO (► **Fig. 2**) and had been carefully monitored. Esophagogastroduodenoscopy on admission revealed the coil extending through the stomach and hooking into the mucosa at the angular portion of the stomach (► **Fig. 3**). Abdominal computed tomography revealed that the coil tip was now in the jejunum (► **Fig. 4**). Because of the risks of ulceration, perforation, or intussusception, we decided to remove it. Oral double-balloon enteroscopy (DBE) detected the coil tip enveloped by food residue in the jejunum; we carefully grasped it using forceps and pulled it back into the stomach. At first, we failed to cut the coil



► **Fig. 2** Endoscopic view of coil used to treat gastric varices, migrating from the fornix of the stomach (arrow).

wire using a scissor-type electrical knife and loop cutter, but we finally succeeded in cutting it using argon plasma coagulation (APC) (► **Video 1**). There were no adverse events during this procedure and the patient's symptoms improved. The recovered coil was an 82-cm cerecyte coil. It was only possible to cut it using APC because the coil wire had unraveled and lengthened (► **Fig. 5**). Compared with radiologists, few gastroenterologists know about migrated PTO coils and their characteristics.

This case shows a rare complication of PTO that was successfully treated by APC and DBE. We propose to describe this "coil valve syndrome" as "ball valve-like syndrome due to deviated coil."

Endoscopy_UCTN_Code_CPL_1AK_2AH



► **Fig. 3** Endoscopic view of a migrated coil biting into the mucous membrane at the angular portion of the stomach.



▶ **Video 1** “Coil valve syndrome”: a rare complication of percutaneous transhepatic obliteration for gastric varices that was successfully treated by argon plasma coagulation (APC) and double-balloon endoscopy (DBE).

Corresponding author

Koji Nagaike, MD

Department of Gastroenterology and Hepatology, Suita Municipal Hospital, 5-7 Kishibeshinmachi, Osaka 564-8567, Japan

Fax: +81-6-63805825

nagaike.koji@gmail.com

References

- [1] Lunderquist A, Vang J. Transhepatic catheterization and obliteration of the coronary vein in patients with portal hypertension and esophageal varices. *N Engl J Med* 1974; 291: 646–649
- [2] Kawai N, Minamiguchi H, Sato M et al. Percutaneous transportal outflow-vessel-occluded sclerotherapy for gastric varices unmanageable by balloon-occluded retrograde transvenous obliteration. *Hepatol Res* 2013; 43: 430–435
- [3] Adebajo CO, Waxman I, Chapman C et al. Foiled by coils: upper GI bleeding from a rare delayed adverse event of transarterial embolization. *Gastrointest Endosc* 2017; 85: 1295–1296
- [4] Hobbs WH, Cohen SE. Gastroduodenal invagination due to a submucous lipoma of the stomach. *Am J Surg* 1946; 71: 505–518

Bibliography

DOI <https://doi.org/10.1055/a-0881-2667>

Published online: 12.4.2019

Endoscopy 2019; 51: E199–E200

© Georg Thieme Verlag KG

Stuttgart · New York

ISSN 0013-726X



▶ **Fig. 4** Computed tomography (CT) confirmed that the end of the coil was in the jejunum.



▶ **Fig. 5** The recovered unraveled 82-cm cerecyte coil.

Competing interests

None

The authors

Koji Nagaike¹, Shiro Hayashi², Kengo Nagai¹, Hirokazu Sasakawa¹, Kiyonori Yuguchi¹, Yuichi Yoshida¹, Masafumi Naito¹

1 Department of Gastroenterology and Hepatology, Suita Municipal Hospital, Osaka, Japan

2 Department of Gastroenterology and Internal Medicine, Hayashi Clinic, Osaka, Japan

CORRECTION

Koji Nagaike, Shiro Hayashi, Kengo Nagai et al. Coil valve syndrome: a rare complication of percutaneous transhepatic obliteration successfully treated by argon plasma coagulation and double-balloon endoscopy.

Endoscopy 2019,

doi:10.1055/a-0881-2667

In the above-mentioned article the name of the author Yuichi Yoshida has been corrected.

This was corrected in the online version on May 27, 2019.