

## Novel mouth guard for safe endoscopy in the COVID era




► **Fig. 1** Attach the upper part of a baby's feeding bottle to the mouth guard.



► **Fig. 2** A silicon nipple with a slit functions as a valve.



► **Fig. 3** Invert the nipple into the bottle neck and screw down the collar.



► **Video 1** Novel mouth guard for safe endoscopy.

COVID-19 is a disease caused by a novel coronavirus (SARS-CoV-2), which is chiefly transmitted through respiratory secretions, aerosols, and contaminated surfaces [1].

Endoscopy is an aerosol-generating procedure and thus carries a significant risk of transmission of aerosol-borne diseases to the endoscopist and staff [2]. During endoscopic procedures, a wide open space is left between the mouth guard and the endoscope, through which large quantities of aerosols escape, increasing the risk of infection.

We have made a few changes in the mouth guard to prevent exposure to aerosols and allow endoscopy to be carried out safely. The elements of the novel mouth guard are: a standard mouth guard, a baby feeding bottle, a feeding bottle nipple, and a triple-layer mask (► **Video 1**).

Take a plastic feeding bottle and cut off the upper end at the neck. Fix this upper end of the bottle to the mouth guard (► **Fig. 1**). Then take the nipple and make a small slit across its top (► **Fig. 2**). Invert the nipple (► **Fig. 3**), place it in the bottle neck, and screw down the collar tightly



► **Fig. 4** Perspective view of the novel mouth guard.

(► **Fig. 4**). Then take a surgical mask, cut a hole in the middle of it, and fix the assembly into the mask (► **Fig. 5**). Place the novel mouth guard inside the patient's mouth and perform endoscopy with all due precautions and necessary protective equipment. The nipple inside the device acts as a perfect valve and prevents aerosols from escaping (► **Video 1**). We have been able to carry out both diagnostic and therapeutic endoscopies with



► **Fig. 5** Novel mouth guard in place in a triple-layer mask, ready for use.

these changes. We have demonstrated the efficacy of our device by performing a simple experiment simulating endoscopy and aerosol generation (► **Video 1**). Our device is cheap, easy to make, reusable, and allows us to do endoscopy more safely during this COVID era.

Endoscopy\_UCTN\_Code\_TTT\_1AO\_2AN

## Acknowledgments

We thank Mr. Kuldeep Singh Katiyar for editing the video.

## Competing interests

The authors declare that they have no conflict of interest.

## The authors

**Sachin Dev Munjal<sup>1</sup>, Yogita Munjal<sup>2</sup>**

- 1 Department of Gastroenterology, Gastro Liver Care Hospital, Saharanpur, Uttar Pradesh, India
- 2 Department of Histopathology, Gastro Liver Care Hospital, Saharanpur, Uttar Pradesh, India

## Corresponding author

**Sachin Dev Munjal, MD**

Department of Gastroenterology, Gastro Liver Care and Saksham Hospital, Saharanpur, 247001 Uttar Pradesh, India  
gastro.liver.help@gmail.com

## References

- [1] del Rio C, Malani PN. COVID-19 – new insights on a rapidly changing epidemic. *JAMA* 2020; 323: 1339–1340
- [2] Johnston ER, Habib-Bein N, Dueker JM. Risk of bacterial exposure to the endoscopist's face during endoscopy. *Gastrointest Endosc* 2019; 89: 818–824

## Bibliography

Endoscopy 2021; 53: E238–E239

DOI 10.1055/a-1308-0778

ISSN 0013-726X

published online 17.12.2020

© 2020. Thieme. All rights reserved.

Georg Thieme Verlag KG, Rüdigerstraße 14, 70469 Stuttgart, Germany

### ENDOSCOPY E-VIDEOS

<https://eref.thieme.de/e-videos>



Endoscopy E-Videos is a free access online section, reporting on interesting cases and new

techniques in gastroenterological endoscopy. All papers include a high quality video and all contributions are freely accessible online.

This section has its own submission website at

<https://mc.manuscriptcentral.com/e-videos>