

Initial trimming followed by complete removal of an esophageal self-expandable metal stent for stent-related symptoms




Fig. 1 Endoscopic images showing: **a** the esophageal covered metal stent with its distal edge protruding into the stomach wall (retrograde view); **b** the esophageal covered metal stent being trimmed using argon plasma coagulation in retroflexed view.



Fig. 2 The transected portion of the covered metal stent.




Fig. 3 The remainder of the covered metal stent following its complete removal 3 months later.



Fig. 4 Esophagogram showing a patent esophagus with no extravasation of contrast.



Endoscopic trimming of the esophageal covered metal stent using argon plasma coagulation: the esophageal covered metal stent is seen protruding into the stomach wall; the distal part of the stent is trimmed using argon plasma coagulation in retroflexed view; the transected stent is removed using a snare.



Endoscopic removal of the esophageal covered metal stent: the distal part of the stent is grabbed with a rat-toothed forceps, and the stent is removed completely using an inversion technique by rotating and withdrawing the endoscope.

Placement of long, protruding self-expandable metal stents (SEMSs) into the gastrointestinal lumen may cause related symptoms. A few reports have described the usefulness of argon plasma coagulation (APC) for trimming or fenestrating a

SEMS [1–4]. We report a trimming technique for a covered SEMS in the esophagus using APC in a retrograde fashion, followed by its complete removal. A 67-year-old woman presented with dysphagia. Esophagogastroduodenoscopy

(EGD) showed a large ulcerated tumor in the esophagus with tumor excavation. A 12-cm partially covered SEMS was placed across the tumor. Subsequently the patient was able to resume eating solid food and underwent chemotherapy. However, 1 month after stent placement, she developed epigastric pain and dysphagia from impaction of the stent into the proximal stomach (▶ Fig. 1 a). The distal portion of the stent was trimmed with APC using a generator at a setting of 80 W and a flow rate of 2 L/min (▶ Fig. 1 b; ▶ Video 1). The procedure was performed with the scope in a retroflexed position to prevent esophageal mucosal injury. A length of the stent (approximately 4 cm) was completely severed in a circumferential manner and was successfully removed from the stomach (▶ Fig. 2). After the procedure, the patient's pain and dysphagia improved.

After 3 months, however, she developed severe acid reflux and we decided to remove the remainder of the stent. Hyperplastic tissue at the uncovered proximal part of the stent was leveled using a stiff snare and APC to free up some of the mesh from the mucosa. The distal part of the stent was then grabbed with a rat-toothed forceps, and the endoscope was withdrawn in a steady rotational fashion, such that the mesh eventually inverted, was dislodged, and then was successfully removed en bloc (▶ Fig. 3; ▶ Video 2). A subsequent esophagogram demonstrated

improvement of the stricture without evidence of contrast extravasation (● Fig. 4). All of the patient's stent-related symptoms resolved after these interventions.

Endoscopy_UCTN_Code_TTT_1AO_2AZ

Competing interests: None

**Takeshi Tsujino, John G. Lee,
Kenneth J. Chang**

Division of Gastroenterology and Hepatology, H.H. Chao Comprehensive Digestive Disease Center, University of California, Orange, California, USA

References

- 1 *Rerknimitr R, Naprasert P, Kongkam P et al.* Trimming a metallic biliary stent using an argon plasma coagulator. *Cardiovasc Intervent Radiol* 2007; 30: 534–536
- 2 *Ishii K, Itoi T, Sofuni A et al.* Endoscopic removal and trimming of distal self-expandable metallic biliary stents. *World J Gastroenterol* 2011; 17: 2652–2657
- 3 *Hamada T, Nakai Y, Isayama H et al.* Trimming a covered metal stent during hepaticogastrostomy by using argon plasma coagulation. *Gastrointest Endosc* 2013; 78: 817
- 4 *Tieu AH, Saxena P, Singh VK et al.* Fenestration of a covered metal stent during cystoduodenostomy using argon plasma coagulation. *Endoscopy* 2014; 46 (Suppl. 01): E512–E513

Bibliography

DOI <http://dx.doi.org/10.1055/s-0042-102881>
Endoscopy 2016; 48: E109–E110
© Georg Thieme Verlag KG
Stuttgart · New York
ISSN 0013-726X

Corresponding author

Kenneth J. Chang, MD
H.H. Chao Comprehensive Digestive Disease Center, University of California
Irvine Medical Center
101 The City Drive, Bldg. 22C
Orange
CA 92868
USA
Fax: +1-714-456-7520
kchang@uci.edu